

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	CTO WE13		Date Received:	
Client Sample ID:	PB165497BL		SDG No.:	P5166
Lab Sample ID:	PB165497BL		Matrix:	Water
Analytical Method:	SW8270SIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3510C			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN035511.D	1	12/09/24 11:27	12/09/24 17:43	PB165497

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.46		30 - 150		114%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.45		30 - 150		112%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.45	*	55 - 111		113%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.43		53 - 106		106%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.56	*	58 - 132		141%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1640	7.3				
1146-65-2	Naphthalene-d8	4580	10.052				
15067-26-2	Acenaphthene-d10	3250	13.957				
1517-22-2	Phenanthrene-d10	7570	16.736				
1719-03-5	Chrysene-d12	6430	20.974				
1520-96-3	Perylene-d12	6250	23.067				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products