

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: SDG No.: MJNL16
Instrument ID: P8 Start Date: 12/13/2024
Analytical Method: ICP-MS Run Batch: LB133950
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	0.0	0	1	1.1	10	50	52	4
Copper	0	0.0	0	2	2.1	5	500	550	10
Lead	0	0.0	0	1	1.0	0	250	260	4
Zinc	0	0.0	0	5	5.3	6	500	520	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: _____ SDG No.: MJNL16
Instrument ID: P8 Start Date: 12/13/2024
Analytical Method: ICP-MS Run Batch: LB133950
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	125	130	4	250	260	4	500	510	2
Copper	1250	1300	4	2500	2700	8	5000	5100	2
Lead	625	650	4	1250	1300	4	2500	2500	0
Zinc	1250	1300	4	2500	2700	8	5000	5200	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: SDG No.: MJNL16
Instrument ID: P8 Start Date: 12/13/2024
Analytical Method: ICP-MS Run Batch: LB133950
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	1000	990	-1						
Copper	10000	9900	-1						
Lead	5000	5000	0						
Zinc	10000	9900	-1						