

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNL16
 Initial Calibration Verification Source : MP83041
 Continuing Calibration Verification Source : MP83042
 Run Batch: LB133950 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV001				ID: CCV001				ID: CCV002		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	200	220	109	1	500	470	95	1	480	97	1
Copper	100	100	103	1	5000	4600	92	1	4600	92	1
Lead	200	210	104	2	2500	2500	99	0	2600	105	9
Zinc	200	210	105	1	5000	5000	100	1	5000	100	0

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNL16
 Initial Calibration Verification Source : MP83041
 Continuing Calibration Verification Source : MP83042
 Run Batch: LB133950 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV003				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic					500	500	100	1			
Copper					5000	4700	94	1			
Lead					2500	2600	103	1			
Zinc					5000	4900	99	1			