

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: _____ SDG No.: MJNLH1
Instrument ID: P5 Start Date: 01/02/2025
Analytical Method: ICP-AES Run Batch: LB134151
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	0	-0	0	10	10	0	1250	1300	4
Barium	0	-0.020	0	200	220	10	2500	2500	0
Cadmium	0	-0	0	5	5.3	6	625	630	1
Chromium	0	-0	0	10	11	10	3875	4100	6
Lead	0	-0	0	10	11	10	6250	6100	-2
Selenium	0	0.010	0	35	31	-11	1250	1200	-4
Silver	0	-0	0	10	10	0	312.5	310	-1

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: SDG No.: MJNLH1
Instrument ID: P5 Start Date: 01/02/2025
Analytical Method: ICP-AES Run Batch: LB134151
Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Arsenic	2500	2600	4	5000	5000	0	10000	9800	-2
Barium	5000	5100	2	10000	10000	0	20000	20000	0
Cadmium	1250	1300	4	2500	2500	0	5000	5000	0
Chromium	7750	8000	3	15500	16000	3	31000	30000	-3
Lead	12500	12000	-4	25000	25000	0	50000	51000	2
Selenium	2500	2500	0	5000	4900	-2	10000	10000	0
Silver	625	630	1	1250	1300	4	2500	2500	0