

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNLH1
 Initial Calibration Verification Source : MP83508
 Continuing Calibration Verification Source : MP83511
 Run Batch: LB134151 Analytical Method: ICP-AES
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV007				ID: CCV055				ID: CCV056		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Arsenic	1000	1100	107	0	5000	4900	99	0	5100	102	0
Barium	520	530	102	0	10000	10000	102	1	10000	104	1
Cadmium	510	510	100	0	2500	2500	99	0	2600	103	0
Chromium	520	570	109	0	15500	16000	101	0	16000	103	0
Lead	1000	990	99	0	25000	25000	99	0	26000	102	0
Selenium	1000	1000	100	1	5000	5000	99	0	5200	103	1
Silver	250	250	102	1	1250	1300	102	1	1300	103	0