

LABORATORY CONTROL SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51821 MA No.: _____ SDG No.: MJNLH1
Matrix : Water Preparation Method: 200.7
Analytical Method: ICP-AES Preparation Batch: PB165673
Concentration Units ($\mu\text{g/L}$, mg/L , mg/kg dry weight, or μg): $\mu\text{g/L}$

Analyte	True	Found	%R
Arsenic	20	25	123
Barium	400	410	102
Cadmium	10	10	100
Chromium	20	21	104
Lead	20	20	100
Selenium	70	65	92
Silver	20	20	99