

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51821 MA No. : SDG No.: MJNLH1
 Initial Calibration Verification Source : MP83660
 Continuing Calibration Verification Source : MP83662
 Run Batch: LB133975 Analytical Method: CVAA
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV021				ID: CCV035				ID: CCV036		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Mercury	4.0	3.9	97		5.0	5.0	101		5.0	100	