

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE002	TUNE	12/18/24 10:50		Jaswal	OK
2	S00	S0	CAL1	12/18/24 13:30		Jaswal	OK
3	S02	S02	CAL3	12/18/24 13:36		Jaswal	OK
4	S03	S03	CAL4	12/18/24 13:40		Jaswal	OK
5	S04	S04	CAL5	12/18/24 13:43		Jaswal	OK
6	S05	S05	CAL6	12/18/24 13:59		Jaswal	OK
7	S06	S06	CAL7	12/18/24 14:02		Jaswal	OK
8	S07	S07	CAL8	12/18/24 14:05		Jaswal	OK
9	S08	S08	CAL9	12/18/24 14:07		Jaswal	OK
10	ICV002	ICV002	ICV	12/18/24 14:28		Jaswal	OK
11	ICB002	ICB002	ICB	12/18/24 14:31		Jaswal	OK
12	ICSA002	ICSA002	ICSA	12/18/24 14:48		Jaswal	OK
13	ICSAB002	ICSAB002	ICSAB	12/18/24 14:58		Jaswal	OK
14	CCV011	CCV011	CCV	12/18/24 15:01		Jaswal	OK
15	CCB011	CCB011	CCB	12/18/24 15:04		Jaswal	OK
16	PB165601BL	PBS601	MB	12/18/24 15:07		Jaswal	OK
17	PB165601BS	LCS601	LCS	12/18/24 15:10		Jaswal	OK
18	PB165601BSD	LCSD601	LCSD	12/18/24 15:14		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

19	PB165617BL	PBS617	MB	12/18/24 15:17		Jaswal	OK
20	PB165617BS	LCS617	LCS	12/18/24 15:20		Jaswal	OK
21	PB165617BSD	LCSD617	LCSD	12/18/24 15:23		Jaswal	OK
22	PB165633BL	PBS633	MB	12/18/24 15:27		Jaswal	OK
23	PB165633BS	LCS633	LCS	12/18/24 15:30		Jaswal	OK
24	PB165633BSD	LCSD563	LCSD	12/18/24 15:33		Jaswal	OK
25	PB165635BL	PBS635	MB	12/18/24 15:36		Jaswal	OK
26	PB165635BS	LCS635	LCS	12/18/24 15:40		Jaswal	OK
27	PB165635BSD	LCSD635	LCSD	12/18/24 15:43		Jaswal	OK
28	PB165657BL	PBS657	MB	12/18/24 15:46		Jaswal	OK
29	PB165657BS	LCS657	LCS	12/18/24 15:50		Jaswal	OK
30	PB165657BSD	LCSD657	LCSD	12/18/24 15:53		Jaswal	OK
31	P5172-20	MYCZ38	SAM	12/18/24 15:56	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
32	P5172-21	MYCZ38D	DUP	12/18/24 15:59	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
33	P5172-20L	MYCZ38L	SD	12/18/24 16:02	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
34	P5172-22	MYCZ38S	MS	12/18/24 16:05	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
35	P5172-20MS	MYCZ38SRE	MS	12/18/24 16:08	Ba,Co,Ni,V High , MS Fail for Pb(228.31ppb),Ag(Below RL),Co(34.48ppb),Ni(75.66ppb),V(118.40ppb)	Jaswal	Dilution
36	P5173-20	MYCZ58	SAM	12/18/24 16:11	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	
Chk Standard	MP83635,MP83634

37	P5173-21	MYCZ58D	DUP	12/18/24 16:14	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
38	P5173-20L	MYCZ58L	SD	12/18/24 16:17	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
39	P5173-22	MYCZ58S	MS	12/18/24 16:20	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
40	P5173-20MS	MYCZ58SRE	MS	12/18/24 16:23	Ba,Co,Ni,V High , MS Fail For Ag(Below RL)	Jaswal	Dilution
41	P5177-20	MYD021	SAM	12/18/24 16:26	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
42	P5177-21	MYD021D	DUP	12/18/24 16:30	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
43	P5177-20L	MYD021L	SD	12/18/24 16:33	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
44	P5177-22	MYD021S	MS	12/18/24 16:36	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
45	P5177-20MS	MYD021SRE	MS	12/18/24 16:39	Ba,Co,Ni,V High	Jaswal	Dilution
46	CCV012	CCV012	CCV	12/18/24 16:42		Jaswal	OK
47	CCB012	CCB012	CCB	12/18/24 16:44		Jaswal	OK
48	P5178-20	MYCZ88	SAM	12/18/24 16:48	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
49	P5178-21	MYCZ88D	DUP	12/18/24 16:51	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
50	P5178-20L	MYCZ88L	SD	12/18/24 16:54	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
51	P5178-22	MYCZ88S	MS	12/18/24 16:57	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
52	P5178-20MS	MYCZ88SRE	MS	12/18/24 17:00	Ba,Co,Ni,V High	Jaswal	Dilution
53	P5179-20	MYCZA8	SAM	12/18/24 17:03	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
54	P5179-21	MYCZA8D	DUP	12/18/24 17:06	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
55	P5179-20L	MYCZA8L	SD	12/18/24 17:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

56	P5179-22	MYCZA8S	MS	12/18/24 17:12	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
57	P5179-20MS	MYCZA8SRE	MS	12/18/24 17:15	Ba,Co,Ni,V High	Jaswal	Dilution
58	P5172-20DL	MYCZ38	SAM	12/18/24 17:18	4X For Ba,Co,Ni,V	Jaswal	Confirms
59	P5172-21DL	MYCZ38D	DUP	12/18/24 17:21	4X For Ba,Co,Ni,V	Jaswal	Confirms
60	P5172-20LDL	MYCZ38L	SD	12/18/24 17:24	20X For Ba,Co,Ni,V	Jaswal	Confirms
61	P5172-22DL	MYCZ38S	MS	12/18/24 17:28	4X For Ba,Co,Ni,V	Jaswal	Confirms
62	P5172-20MSDL	MYCZ38SRE	MS	12/18/24 17:31	4X For Ba,Co,Ni,V	Jaswal	Confirms
63	P5173-20DL	MYCZ58	SAM	12/18/24 17:34	4X For Ba,Co,Ni,V	Jaswal	Confirms
64	P5173-21DL	MYCZ58D	DUP	12/18/24 17:37	4X For Ba,Co,Ni,V	Jaswal	Confirms
65	P5173-20LDL	MYCZ58L	SD	12/18/24 17:40	20X For Ba,Co,Ni,V	Jaswal	Confirms
66	P5173-22DL	MYCZ58S	MS	12/18/24 17:43	4X For Ba,Co,Ni,V	Jaswal	Confirms
67	P5173-20MSDL	MYCZ58SRE	MS	12/18/24 17:46	4X For Ba,Co,Ni,V	Jaswal	Confirms
68	CCV013	CCV013	CCV	12/18/24 17:54		Jaswal	OK
69	CCB013	CCB013	CCB	12/18/24 17:57		Jaswal	OK
70	P5177-20DL	MYD021	SAM	12/18/24 18:04	4X For Ba,Co,Ni,V	Jaswal	Confirms
71	P5177-21DL	MYD021D	DUP	12/18/24 18:07	4X For Ba,Co,Ni,V	Jaswal	Confirms
72	P5177-20LDL	MYD021L	SD	12/18/24 18:10	20X For Ba,Co,Ni,V	Jaswal	Confirms
73	P5177-22DL	MYD021S	MS	12/18/24 18:13	4X For Ba,Co,Ni,V	Jaswal	Confirms
74	P5177-20MSDL	MYD021SRE	MS	12/18/24 18:17	4X For Ba,Co,Ni,V	Jaswal	Confirms
75	P5178-20DL	MYCZ88	SAM	12/18/24 18:20	4X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

76	P5178-21DL	MYCZ88D	DUP	12/18/24 18:23	4X For Ba,Co,Ni,V	Jaswal	Confirms
77	P5178-20LDL	MYCZ88L	SD	12/18/24 18:26	20X For Ba,Co,Ni,V	Jaswal	Confirms
78	P5178-22DL	MYCZ88S	MS	12/18/24 18:29	4X For Ba,Co,Ni,V	Jaswal	Confirms
79	P5178-20MSDL	MYCZ88SRE	MS	12/18/24 18:32	4X For Ba,Co,Ni,V	Jaswal	Confirms
80	P5179-20DL	MYCZA8	SAM	12/18/24 18:35	4X For Ba,Co,Ni,V	Jaswal	Confirms
81	P5179-21DL	MYCZA8D	DUP	12/18/24 18:38	4X For Ba,Co,Ni,V	Jaswal	Confirms
82	P5179-20LDL	MYCZA8L	SD	12/18/24 18:42	20X For Ba,Co,Ni,V	Jaswal	Confirms
83	P5179-22DL	MYCZA8S	MS	12/18/24 18:45	4X For Ba,Co,Ni,V	Jaswal	Confirms
84	P5179-20MSDL	MYCZA8SRE	MS	12/18/24 18:48	4X For Ba,Co,Ni,V	Jaswal	Confirms
85	P5184-01	MYCZD9	SAM	12/18/24 18:51		Jaswal	OK
86	P5184-02	MYCZE0	SAM	12/18/24 18:54		Jaswal	OK
87	P5184-03	MYCZE1	SAM	12/18/24 18:57		Jaswal	OK
88	P5184-04	MYCZE2	SAM	12/18/24 19:00		Jaswal	OK
89	P5184-05	MYCZE3	SAM	12/18/24 19:03		Jaswal	OK
90	P5184-06	MYCZE4	SAM	12/18/24 19:06		Jaswal	OK
91	P5184-07	MYCZE5	SAM	12/18/24 19:09		Jaswal	OK
92	CCV014	CCV014	CCV	12/18/24 19:18		Jaswal	OK
93	CCB014	CCB014	CCB	12/18/24 19:21		Jaswal	OK
94	P5184-08	MYCZE6	SAM	12/18/24 19:24		Jaswal	OK
95	P5184-09	MYCZE7	SAM	12/18/24 19:27		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

96	P5184-10	MYCZE8	SAM	12/18/24 19:30		Jaswal	OK
97	P5184-11	MYCZE9	SAM	12/18/24 19:33		Jaswal	OK
98	P5184-12	MYCZF0	SAM	12/18/24 19:36		Jaswal	OK
99	P5184-13	MYCZF1	SAM	12/18/24 19:39		Jaswal	OK
100	P5184-14	MYCZF2	SAM	12/18/24 19:42		Jaswal	OK
101	P5184-15	MYCZF3	SAM	12/18/24 19:45		Jaswal	OK
102	P5184-16	MYCZF4	SAM	12/18/24 19:48		Jaswal	OK
103	P5184-17	MYCZF5	SAM	12/18/24 19:51		Jaswal	OK
104	P5184-18	MYCZF6	SAM	12/18/24 19:54		Jaswal	OK
105	P5184-19	MYCZF7	SAM	12/18/24 19:57		Jaswal	OK
106	P5184-20	MYCZF8	SAM	12/18/24 20:01	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
107	P5184-21	MYCZF8D	DUP	12/18/24 20:04	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
108	P5184-20L	MYCZF8L	SD	12/18/24 20:07	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
109	P5184-22	MYCZF8S	MS	12/18/24 20:10	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
110	P5184-20MS	MYCZF8SRE	MS	12/18/24 20:13	Ba,Co,Ni,V High , - MS Fail For Sb,Be,Se,Tl(Below RL),Ag(1.21ppb)	Jaswal	Dilution
111	P5186-01	MYCZN9	SAM	12/18/24 20:15		Jaswal	OK
112	P5186-02	MYCZP0	SAM	12/18/24 20:19		Jaswal	OK
113	P5186-03	MYCZP1	SAM	12/18/24 20:22		Jaswal	OK
114	P5186-04	MYCZP2	SAM	12/18/24 20:25		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

115	P5186-05	MYCZP3	SAM	12/18/24 20:28		Jaswal	OK
116	P5186-06	MYCZP4	SAM	12/18/24 20:31		Jaswal	OK
117	P5186-07	MYCZP5	SAM	12/18/24 20:34		Jaswal	OK
118	P5186-08	MYCZP6	SAM	12/18/24 20:37		Jaswal	OK
119	P5186-09	MYCZP7	SAM	12/18/24 20:40		Jaswal	OK
120	CCV015	CCV015	CCV	12/18/24 20:43		Jaswal	OK
121	CCB015	CCB015	CCB	12/18/24 20:50		Jaswal	OK
122	P5186-10	MYCZP8	SAM	12/18/24 20:53		Jaswal	OK
123	P5186-11	MYCZP9	SAM	12/18/24 20:56		Jaswal	OK
124	P5186-12	MYCZQ0	SAM	12/18/24 20:59		Jaswal	OK
125	P5186-13	MYCZQ1	SAM	12/18/24 21:02		Jaswal	OK
126	P5186-14	MYCZQ2	SAM	12/18/24 21:05		Jaswal	OK
127	P5186-15	MYCZQ3	SAM	12/18/24 21:08		Jaswal	OK
128	P5186-16	MYCZQ4	SAM	12/18/24 21:11		Jaswal	OK
129	P5186-17	MYCZQ5	SAM	12/18/24 21:14		Jaswal	OK
130	P5186-18	MYCZQ6	SAM	12/18/24 21:18		Jaswal	OK
131	P5186-19	MYCZQ7	SAM	12/18/24 21:21		Jaswal	OK
132	P5186-20	MYCZQ8	SAM	12/18/24 21:24	Ba,Co,NI,V High in MSRE	Jaswal	Dilution
133	P5186-21	MYCZQ8D	DUP	12/18/24 21:27	Ba,Co,NI,V High in MSRE	Jaswal	Dilution
134	P5186-20L	MYCZQ8L	SD	12/18/24 21:30	Ba,Co,NI,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	
Chk Standard	MP83635,MP83634

135	P5186-22	MYCZQ8S	MS	12/18/24 21:33	Ba,Co,NI,V High in MSRE	Jaswal	Dilution
136	P5186-20MS	MYCZQ8SRE	MS	12/18/24 21:36	Ba,Co,NI,V High , MS Fail For Ag(Below RL),Ni(71.24ppb),V(189.17)	Jaswal	Dilution
137	P5180-01	MYCYW9	SAM	12/18/24 21:39		Jaswal	OK
138	P5180-02	MYCYX0	SAM	12/18/24 21:42		Jaswal	OK
139	P5180-03	MYCYX1	SAM	12/18/24 21:45		Jaswal	OK
140	P5180-04	MYCYX2	SAM	12/18/24 21:48		Jaswal	OK
141	P5180-05	MYCYX3	SAM	12/18/24 21:51		Jaswal	OK
142	P5180-06	MYCYX4	SAM	12/18/24 21:54		Jaswal	OK
143	P5180-07	MYCYX5	SAM	12/18/24 21:57		Jaswal	OK
144	P5180-08	MYCYX6	SAM	12/18/24 22:01		Jaswal	OK
145	P5180-09	MYCYX7	SAM	12/18/24 22:04		Jaswal	OK
146	P5180-10	MYCYX8	SAM	12/18/24 22:07		Jaswal	OK
147	CCV016	CCV016	CCV	12/18/24 22:10		Jaswal	OK
148	CCB016	CCB016	CCB	12/18/24 22:12		Jaswal	OK
149	P5180-11	MYCZA9	SAM	12/18/24 22:16		Jaswal	OK
150	P5180-12	MYCZB0	SAM	12/18/24 22:19		Jaswal	OK
151	P5180-13	MYCZB1	SAM	12/18/24 22:22		Jaswal	OK
152	P5180-14	MYCZB2	SAM	12/18/24 22:25		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

153	P5180-15	MYCZB3	SAM	12/18/24 22:28		Jaswal	OK
154	P5180-16	MYCZB4	SAM	12/18/24 22:32		Jaswal	OK
155	P5180-17	MYCZB5	SAM	12/18/24 22:35		Jaswal	OK
156	P5180-18	MYCZB6	SAM	12/18/24 22:38		Jaswal	OK
157	P5180-19	MYCZB7	SAM	12/18/24 22:41		Jaswal	OK
158	P5180-20	MYCZB8	SAM	12/18/24 22:44	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
159	P5180-21	MYCZB8D	DUP	12/18/24 22:47	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
160	P5180-20L	MYCZB8L	SD	12/18/24 22:50	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
161	P5180-22	MYCZB8S	MS	12/18/24 22:54	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
162	P5180-20MS	MYCZB8SRE	MS	12/18/24 22:57	Ba,Co,Ni,V High	Jaswal	Dilution
163	P5184-20DL	MYCZF8	SAM	12/18/24 22:59	4X For Ba,Co,Ni,V	Jaswal	Confirms
164	P5184-21DL	MYCZF8D	DUP	12/18/24 23:03	4X For Ba,Co,Ni,V	Jaswal	Confirms
165	P5184-20LDL	MYCZF8L	SD	12/18/24 23:06	20X For Ba,Co,Ni,V	Jaswal	Confirms
166	P5184-22DL	MYCZF8S	MS	12/18/24 23:09	4X For Ba,Co,Ni,V	Jaswal	Confirms
167	P5184-20MSDL	MYCZF8SRE	MS	12/18/24 23:12	4X For Ba,Co,Ni,V	Jaswal	Confirms
168	P5186-20DL	MYCZQ8	SAM	12/18/24 23:15	4X For Ba,Co,Ni,V	Jaswal	Confirms
169	P5186-21DL	MYCZQ8D	DUP	12/18/24 23:18	4X For Ba,Co,Ni,V	Jaswal	Confirms
170	P5186-20LDL	MYCZQ8L	SD	12/18/24 23:22	20X For Ba,Co,Ni,V	Jaswal	Confirms
171	P5186-22DLRE	MYCZQ8S	MS	12/18/24 23:25	Not Use	Jaswal	Not Ok
172	P5186-20MSDL	MYCZQ8SRE	MS	12/18/24 23:28	4X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	
LCS Standard	
Chk Standard	MP83635,MP83634

173	P5180-20DL	MYCZB8	SAM	12/18/24 23:31	4X For Ba,Co,Ni,V	Jaswal	Confirms
174	P5180-21DL	MYCZB8D	DUP	12/18/24 23:34	4X For Ba,Co,Ni,V	Jaswal	Confirms
175	CCV017	CCV017	CCV	12/18/24 23:37		Jaswal	OK
176	CCB017	CCB017	CCB	12/18/24 23:40		Jaswal	OK
177	P5180-20LDL	MYCZB8L	SD	12/18/24 23:43	20X For Ba,Co,Ni,V	Jaswal	Confirms
178	P5180-22DL	MYCZB8S	MS	12/18/24 23:47	4X For Ba,Co,Ni,V	Jaswal	Confirms
179	P5180-20MSDL	MYCZB8SRE	MS	12/18/24 23:50	4X For Ba,Co,Ni,V	Jaswal	Confirms
180	P5180-01DL	MYCYW9	SAM	12/18/24 23:53	Not Use	Jaswal	Not Ok
181	P5180-03DL	MYCYX1	SAM	12/18/24 23:56	Not Use	Jaswal	Not Ok
182	P5180-04DL	MYCYX2	SAM	12/19/24	Not Use	Jaswal	Not Ok
183	P5180-05DL	MYCYX3	SAM	12/19/24 00:03	Not Use	Jaswal	Not Ok
184	P5180-06DL	MYCYX4	SAM	12/19/24 00:06	Not Use	Jaswal	Not Ok
185	P5180-07DL	MYCYX5	SAM	12/19/24 00:09	Not Use	Jaswal	Not Ok
186	P5180-08DL	MYCYX6	SAM	12/19/24 00:12	Not Use	Jaswal	Not Ok
187	P5180-09DL	MYCYX7	SAM	12/19/24 00:15	Not Use	Jaswal	Not Ok
188	P5180-10DL	MYCYX8	SAM	12/19/24 00:18	Not Use	Jaswal	Not Ok
189	P5186-09DL	MYCZP7	SAM	12/19/24 00:22	Not Use	Jaswal	Not Ok
190	P5186-22DL	MYCZQ8S	MS	12/19/24 00:25	4X For Ba,Co,Ni,V	Jaswal	Confirms
191	P5180-10DL2	MYCYX8	SAM	12/19/24 00:49	Not Use	Jaswal	Not Ok
192	CCV018	CCV018	CCV	12/19/24 00:54		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134003

Review By	Sarabjit Jaswal	Review On	12/19/2024 8:41:35 AM
Supervise By	Mohan Bera	Supervise On	12/19/2024 8:42:22 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard			
LCS Standard			
Chk Standard	MP83635,MP83634		

193	CCB018	CCB018	CCB	12/19/24 00:56		Jaswal	OK
-----	--------	--------	-----	----------------	--	--------	----