

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134058

Review By	Sarabjit Jaswal	Review On	12/23/2024 4:13:28 AM
Supervise By	Mohan Bera	Supervise On	12/23/2024 4:14:06 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard	MP83713		
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	TUNE	TUNE004	TUNE	12/20/24 12:30		Jaswal	OK
2	S00	S0	CAL1	12/20/24 13:03		Jaswal	OK
3	S02	S02	CAL3	12/20/24 13:10		Jaswal	OK
4	S03	S03	CAL4	12/20/24 13:13		Jaswal	OK
5	S04	S04	CAL5	12/20/24 13:16		Jaswal	OK
6	S05	S05	CAL6	12/20/24 13:19		Jaswal	OK
7	S06	S06	CAL7	12/20/24 13:22		Jaswal	OK
8	S07	S07	CAL8	12/20/24 13:24		Jaswal	OK
9	S08	S08	CAL9	12/20/24 13:27		Jaswal	OK
10	ICV004	ICV004	ICV	12/20/24 13:41	Inst. paused for Calibration review	Jaswal	OK
11	ICB004	ICB004	ICB	12/20/24 14:31	Inst. paused for review	Jaswal	OK
12	ICSA004	ICSA004	ICSA	12/20/24 14:35		Jaswal	OK
13	ICSAB004	ICSAB004	ICSAB	12/20/24 14:38		Jaswal	OK
14	CCV030	CCV030	CCV	12/20/24 14:41		Jaswal	OK
15	CCB030	CCB030	CCB	12/20/24 14:47	Inst. paused for review	Jaswal	OK
16	PB165687BL	PBS687	MB	12/20/24 14:52		Jaswal	OK
17	PB165687BS	LCS687	LCS	12/20/24 14:55		Jaswal	OK
18	PB165687BSD	LCSD687	LCSD	12/20/24 14:59		Jaswal	OK

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134058

Review By	Sarabjit Jaswal	Review On	12/23/2024 4:13:28 AM
Supervise By	Mohan Bera	Supervise On	12/23/2024 4:14:06 AM

STD. NAME	STD REF.#
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636
ICV Standard	MP83629
CCV Standard	MP83632
ICSA Standard	MP83630,MP83631
CRI Standard	MP83713
LCS Standard	MP83669
Chk Standard	MP83635,MP83634

19	PB165689BL	PBS689	MB	12/20/24 15:02		Jaswal	OK
20	PB165689BS	LCS689	LCS	12/20/24 15:05		Jaswal	OK
21	PB165689BSD	LCSD689	LCSD	12/20/24 15:09		Jaswal	OK
22	PB165692BL	PBS692	MB	12/20/24 15:12		Jaswal	OK
23	PB165692BS	LCS692	LCS	12/20/24 15:15		Jaswal	OK
24	PB165692BSD	LCSD692	LCSD	12/20/24 15:18		Jaswal	OK
25	PB165715BL	PBS715	MB	12/20/24 15:22		Jaswal	OK
26	PB165715BS	LCS715	LCS	12/20/24 15:25		Jaswal	OK
27	PB165715BSD	LCSD715	LCSD	12/20/24 15:28		Jaswal	OK
28	P5188-20	MYCZR8	SAM	12/20/24 15:32	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
29	P5188-21	MYCZR8D	DUP	12/20/24 15:35	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
30	P5188-20L	MYCZR8L	SD	12/20/24 15:38	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
31	P5188-22	MYCZR8S	MS	12/20/24 15:41	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
32	P5188-20MS	MYCZR8SRE	MS	12/20/24 15:44	Ba,Co,Ni,V High	Jaswal	Dilution
33	P5189-20	MYCZK8	SAM	12/20/24 15:47	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
34	P5189-21	MYCZK8D	DUP	12/20/24 15:50	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
35	P5189-20L	MYCZK8L	SD	12/20/24 15:53	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
36	P5189-22	MYCZK8S	MS	12/20/24 15:56	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
37	P5189-20MS	MYCZK8SRE	MS	12/20/24 15:59	Ba,Co,Ni,V High	Jaswal	Dilution
38	P5190-20	MYCZM8	SAM	12/20/24 16:02	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134058

Review By	Sarabjit Jaswal	Review On	12/23/2024 4:13:28 AM
Supervise By	Mohan Bera	Supervise On	12/23/2024 4:14:06 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard	MP83713		
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

39	P5190-21	MYCZM8D	DUP	12/20/24 16:05	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
40	P5190-20L	MYCZM8L	SD	12/20/24 16:09	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
41	P5190-22	MYCZM8S	MS	12/20/24 16:12	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
42	P5190-20MS	MYCZM8SRE	MS	12/20/24 16:15	Ba,Co,Ni,V High	Jaswal	Dilution
43	CCV031	CCV031	CCV	12/20/24 16:18		Jaswal	OK
44	CCB031	CCB031	CCB	12/20/24 16:24	Inst. paused for review	Jaswal	OK
45	P5191-13	MYD033	SAM	12/20/24 16:27	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
46	P5191-14	MYD033D	DUP	12/20/24 16:30	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
47	P5191-13L	MYD033L	SD	12/20/24 16:33	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
48	P5191-15	MYD033S	MS	12/20/24 16:37	Ba,Co,Ni,V High in MSRE	Jaswal	Dilution
49	P5191-13MS	MYD033SRE	MS	12/20/24 16:40	Ba,Co,Ni,V High , MS Fail For As(11.82ppb)	Jaswal	Dilution
50	P5188-20DL	MYCZR8	SAM	12/20/24 16:43	4X For Ba,Co,Ni,V	Jaswal	Confirms
51	P5188-21DL	MYCZR8D	DUP	12/20/24 16:46	4X For Ba,Co,Ni,V	Jaswal	Confirms
52	P5188-20LDL	MYCZR8L	SD	12/20/24 16:49	20X For Ba,Co,Ni,V	Jaswal	Confirms
53	P5188-22DL	MYCZR8S	MS	12/20/24 16:52	4X For Ba,Co,Ni,V	Jaswal	Confirms
54	P5188-20MSDL	MYCZR8SRE	MS	12/20/24 16:55	4X For Ba,Co,Ni,V	Jaswal	Confirms
55	P5189-20DL	MYCZK8	SAM	12/20/24 16:58	4X For Ba,Co,Ni,V	Jaswal	Confirms
56	P5189-21DL	MYCZK8D	DUP	12/20/24 17:02	4X For Ba,Co,Ni,V	Jaswal	Confirms
57	P5189-20LDL	MYCZK8L	SD	12/20/24 17:05	20X For Ba,Co,Ni,V	Jaswal	Confirms

Instrument ID: P7

Daily Analysis Runlog For Sequence/QC Batch ID # LB134058

Review By	Sarabjit Jaswal	Review On	12/23/2024 4:13:28 AM
Supervise By	Mohan Bera	Supervise On	12/23/2024 4:14:06 AM
STD. NAME	STD REF.#		
ICAL Standard	MP83619,MP83628,MP83627,MP83625,MP83623,MP83622,MP83621,MP83620,MP83636		
ICV Standard	MP83629		
CCV Standard	MP83632		
ICSA Standard	MP83630,MP83631		
CRI Standard	MP83713		
LCS Standard	MP83669		
Chk Standard	MP83635,MP83634		

58	P5189-22DL	MYCZK8S	MS	12/20/24 17:08	4X For Ba,Co,Ni,V	Jaswal	Confirms
59	P5189-20MSDL	MYCZK8SRE	MS	12/20/24 17:11	4X For Ba,Co,Ni,V	Jaswal	Confirms
60	P5190-20DL	MYCZM8	SAM	12/20/24 17:14	4X For Ba,Co,Ni,V	Jaswal	Confirms
61	P5190-21DL	MYCZM8D	DUP	12/20/24 17:17	4X For Ba,Co,Ni,V	Jaswal	Confirms
62	P5190-20LDL	MYCZM8L	SD	12/20/24 17:20	20X For Ba,Co,Ni,V	Jaswal	Confirms
63	P5190-22DL	MYCZM8S	MS	12/20/24 17:24	4X For Ba,Co,Ni,V	Jaswal	Confirms
64	P5190-20MSDL	MYCZM8SRE	MS	12/20/24 17:27	4X For Ba,Co,Ni,V	Jaswal	Confirms
65	P5191-13DL	MYD033	SAM	12/20/24 17:30	4X For Ba,Co,Ni,V	Jaswal	Confirms
66	P5191-14DL	MYD033D	DUP	12/20/24 17:33	4X For Ba,Co,Ni,V	Jaswal	Confirms
67	P5191-13LDL	MYD033L	SD	12/20/24 17:36	20X For Ba,Co,Ni,V	Jaswal	Confirms
68	P5191-15DL	MYD033S	MS	12/20/24 17:39	4X For Ba,Co,Ni,V	Jaswal	Confirms
69	P5191-13MSDL	MYD033SRE	MS	12/20/24 17:43	4X For Ba,Co,Ni,V	Jaswal	Confirms
70	P5182-20A	MYCZ18ARE	PS	12/20/24 17:46	PS for Ag 0.02 ml mp83713-10ml sample	Jaswal	OK
71	P5183-20A	MYCZD8ARE	PS	12/20/24 17:49	PS for Ag 0.02 ml mp83713-10ml sample	Jaswal	OK
72	CCV032	CCV032	CCV	12/20/24 17:52		Jaswal	OK
73	CCB032	CCB032	CCB	12/20/24 17:59	Inst. paused for review	Jaswal	OK