

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51779MA No.: _____ SDG No.: MYCZD9Instrument ID: P7Start Date : 12/18/2024Analytical Method: ICP-MSRun Batch : LB134003

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999992	0.000480	0.000259	Lin. Reg	NONE
Antimony	0.999927	0.000811	0.000001	Lin. Reg	NONE
Arsenic	0.999862	0.000161	0.000002	Lin. Reg	NONE
Barium	0.999983	0.000354	0.000003	Lin. Reg	NONE
Beryllium	0.999930	0.000837	0.000018	Lin. Reg	NONE
Cadmium	0.999831	0.000203	0.000168	Lin. Reg	NONE
Calcium	0.999998	0.000078	0.001150	Lin. Reg	NONE
Chromium	0.999951	0.008394	0.001593	Lin. Reg	NONE
Cobalt	0.999945	0.013175	0.000158	Lin. Reg	NONE
Copper	0.999871	0.009576	0.001251	Lin. Reg	NONE
Iron	0.999980	0.000194	0.002111	Lin. Reg	NONE
Lead	0.999978	0.004266	0.000159	Lin. Reg	NONE
Magnesium	0.999978	0.000745	0.000746	Lin. Reg	NONE
Manganese	0.999949	0.006933	0.000767	Lin. Reg	NONE
Molybdenum	0.999996	0.000708	0.000009	Lin. Reg	NONE
Nickel	0.999966	0.003329	0.000444	Lin. Reg	NONE
Potassium	0.999999	0.000987	0.091461	Lin. Reg	NONE
Selenium	0.999861	0.000013	0.000014	Lin. Reg	NONE
Silver	0.999796	0.001013	0.000012	Lin. Reg	NONE
Sodium	0.998444	0.000945	0.029415	Lin. Reg	NONE
Strontium	0.999981	0.000096	0.000014	Lin. Reg	NONE
Thallium	0.999991	0.003201	0.000031	Lin. Reg	NONE
Titanium	0.999975	0.000231	0.000014	Lin. Reg	NONE
Uranium	0.999980	0.004118	0.000002	Lin. Reg	NONE
Vanadium	0.999980	0.007506	0.000022	Lin. Reg	NONE
Zinc	0.999778	0.001629	0.000518	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51779MA No.: _____ SDG No.: MYCZD9Instrument ID: P7Start Date : 12/23/2024Analytical Method: ICP-MSRun Batch : LB134068

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999965	0.000528	0.000255	Lin. Reg	NONE
Antimony	0.999912	0.000864	0.000002	Lin. Reg	NONE
Arsenic	0.999957	0.000175	0.000003	Lin. Reg	NONE
Barium	0.999928	0.000379	0.000008	Lin. Reg	NONE
Beryllium	0.999929	0.000875	0.000033	Lin. Reg	NONE
Cadmium	0.999958	0.000217	0.000173	Lin. Reg	NONE
Calcium	0.999979	0.000083	0.001759	Lin. Reg	NONE
Chromium	0.999993	0.009170	0.002647	Lin. Reg	NONE
Cobalt	0.999988	0.014569	0.000190	Lin. Reg	NONE
Copper	0.999954	0.010523	0.001238	Lin. Reg	NONE
Iron	0.999986	0.000210	0.001548	Lin. Reg	NONE
Lead	0.999952	0.004707	0.000165	Lin. Reg	NONE
Magnesium	0.999878	0.000866	0.000983	Lin. Reg	NONE
Manganese	0.999993	0.007413	0.000443	Lin. Reg	NONE
Molybdenum	0.999952	0.000763	0.000018	Lin. Reg	NONE
Nickel	0.999900	0.003658	0.000673	Lin. Reg	NONE
Potassium	0.999989	0.001058	0.086952	Lin. Reg	NONE
Selenium	0.999786	0.000014	0.000013	Lin. Reg	NONE
Silver	0.999586	0.001064	0.000006	Lin. Reg	NONE
Sodium	0.999996	0.001350	0.037496	Lin. Reg	NONE
Strontium	0.999896	0.000105	0.000025	Lin. Reg	NONE
Thallium	0.999933	0.003513	0.000082	Lin. Reg	NONE
Titanium	0.999965	0.000251	0.000015	Lin. Reg	NONE
Uranium	0.999810	0.004841	0.000001	Lin. Reg	NONE
Vanadium	0.999992	0.008148	0.000019	Lin. Reg	NONE
Zinc	0.999913	0.001754	0.000415	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51779 MA No.: SDG No.: MYCZD9
Instrument ID: P7 Start Date : 12/23/2024
Analytical Method: ICP-MS Run Batch : LB134068

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------