

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZD9
 Initial Calibration Verification Source : MP83629
 Continuing Calibration Verification Source : MP83632
 Run Batch: LB134003 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV002				ID: CCV011				ID: CCV012		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	210	103	2	500	490	97	2	480	97	4
Arsenic	200	210	105	1	500	510	103	1	520	104	1
Barium	100	100	100	1	2500	2500	98	2	2500	98	4
Beryllium	100	100	100	1	500	470	93	2	470	93	4
Cadmium	100	110	106	1	500	510	102	1	510	102	5
Chromium	100	100	101	0	500	500	100	1	500	100	0
Cobalt	100	100	101	0	500	490	98	1	490	98	0
Copper	100	96	96	0	5000	4900	99	0	5000	99	1
Lead	200	190	97	1	2500	2500	100	2	2500	99	4
Nickel	110	100	94	1	500	470	93	0	470	94	0
Selenium	200	210	106	1	500	500	100	2	500	100	5
Silver	50.0	52	104	2	500	500	100	2	500	101	4
Thallium	210	190	92	2	500	500	101	3	500	100	4
Vanadium	100	97	97	0	500	500	100	1	510	102	1
Zinc	200	200	99	1	5000	4900	97	1	4900	98	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZD9
 Initial Calibration Verification Source : MP83629
 Continuing Calibration Verification Source : MP83632
 Run Batch: LB134003 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV013				ID: CCV014		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	490	98	1	500	100	1
Arsenic					500	510	102	0	510	101	1
Barium					2500	2500	99	1	2500	101	1
Beryllium					500	480	97	1	490	97	2
Cadmium					500	510	103	1	530	105	0
Chromium					500	500	100	1	510	102	1
Cobalt					500	490	98	0	500	99	0
Copper					5000	4900	99	0	4900	99	1
Lead					2500	2500	101	1	2600	103	1
Nickel					500	470	93	0	470	95	0
Selenium					500	500	101	1	490	97	2
Silver					500	510	101	0	500	101	1
Thallium					500	510	101	1	520	104	1
Vanadium					500	510	101	0	510	102	1
Zinc					5000	4900	98	1	4800	97	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZD9Initial Calibration Verification Source : MP83629Continuing Calibration Verification Source : MP83632Run Batch: LB134003 Analytical Method: ICP-MSConcentration Units: $\mu\text{g/L}$

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV015				ID: CCV016		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	500	99	0	490	98	2
Arsenic					500	510	102	0	490	99	0
Barium					2500	2500	101	1	2500	99	2
Beryllium					500	480	97	1	480	95	2
Cadmium					500	520	104	1	500	100	1
Chromium					500	510	102	1	500	100	0
Cobalt					500	500	100	1	500	99	1
Copper					5000	4900	99	1	4900	97	1
Lead					2500	2500	102	0	2500	101	3
Nickel					500	470	94	1	470	94	1
Selenium					500	490	97	1	480	96	3
Silver					500	500	100	1	490	97	1
Thallium					500	520	103	1	510	103	4
Vanadium					500	510	101	1	510	102	0
Zinc					5000	4900	97	0	4800	96	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZD9Initial Calibration Verification Source : MP83629Continuing Calibration Verification Source : MP83632Run Batch: LB134003 Analytical Method: ICP-MSConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV017				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	470	94	9			
Arsenic					500	490	99	0			
Barium					2500	2400	96	9			
Beryllium					500	460	91	8			
Cadmium					500	480	95	8			
Chromium					500	500	101	1			
Cobalt					500	500	99	1			
Copper					5000	4900	97	1			
Lead					2500	2400	97	8			
Nickel					500	470	94	1			
Selenium					500	460	92	9			
Silver					500	470	94	8			
Thallium					500	490	99	7			
Vanadium					500	510	101	1			
Zinc					5000	4800	96	1			

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZD9
 Initial Calibration Verification Source : MP83629
 Continuing Calibration Verification Source : MP83632
 Run Batch: LB134068 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV007				ID: CCV041				ID: CCV042		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony	200	210	106	1	500	500	99	6	510	102	1
Arsenic	200	220	109	1	500	500	101	0	490	99	1
Barium	100	100	102	0	2500	2500	102	6	2600	104	1
Beryllium	100	110	107	2	500	510	101	6	500	101	2
Cadmium	100	110	108	1	500	470	94	6	480	96	1
Chromium	100	100	104	1	500	510	101	0	510	101	1
Cobalt	100	100	102	0	500	490	98	1	490	98	1
Copper	100	97	97	0	5000	4800	96	1	4900	97	1
Lead	200	190	97	0	2500	2500	102	6	2600	102	1
Nickel	110	100	95	0	500	470	94	1	470	94	1
Selenium	200	220	109	1	500	480	96	6	490	97	2
Silver	50.0	52	103	1	500	490	98	5	500	100	2
Thallium	210	200	94	0	500	510	102	5	510	102	1
Vanadium	100	100	101	0	500	510	102	0	520	103	1
Zinc	200	200	99	1	5000	4700	95	1	4800	96	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZD9
 Initial Calibration Verification Source : MP83629
 Continuing Calibration Verification Source : MP83632
 Run Batch: LB134068 Analytical Method: ICP-MS
 Concentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV043				ID: CCV044		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Antimony					500	510	101	1	490	98	3
Arsenic					500	500	101	0	480	97	1
Barium					2500	2600	104	1	2500	100	3
Beryllium					500	500	99	1	480	95	1
Cadmium					500	480	96	0	460	92	3
Chromium					500	510	103	1	490	99	1
Cobalt					500	500	100	0	480	97	0
Copper					5000	4900	97	0	4700	94	0
Lead					2500	2600	103	1	2500	100	3
Nickel					500	470	94	0	460	92	1
Selenium					500	480	97	1	470	93	3
Silver					500	500	100	1	480	96	1
Thallium					500	520	103	1	500	100	4
Vanadium					500	520	104	1	500	101	0
Zinc					5000	4800	96	0	4600	93	0