

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No.: _____ SDG No.: MYCPM3
 Instrument ID: P7 Start Date: 12/19/2024
 Analytical Method: ICP-MS Run Batch: LB134025
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	0	0.0	0	20	18	-10	1000	980	-2
Antimony	0	0.0	0	2	2.2	10	50	53	6
Arsenic	0	0.0	0	1	1.2	20	50	54	8
Barium	0	0.0	0	10	11	10	250	260	4
Beryllium	0	0.0	0	1	1.0	0	50	51	2
Cadmium	0	0.0	0	1	1.2	20	50	56	12
Calcium	0	0.0	0	500	550	10	5000	5200	4
Chromium	0	0.0	0	2	2.1	5	50	51	2
Cobalt	0	0.0	0	1	1.1	10	50	51	2
Copper	0	0.0	0	2	1.9	-5	500	530	6
Iron	0	0.0	0	50	54	8	2500	2700	8
Lead	0	0.0	0	1	0.88	-12	250	250	0
Magnesium	0	0.0	0	500	530	6	5000	5000	0
Manganese	0	0.0	0	1	0.88	-12	500	500	0
Nickel	0	0.0	0	1	1.1	10	50	52	4
Potassium	0	0.0	0	500	540	8	2500	2600	4
Selenium	0	0.0	0	5	5.0	0	50	53	6
Silver	0	0.0	0	1	1.1	10	50	57	14
Sodium	0	0.0	0	500	490	-2	5000	5000	0
Thallium	0	0.0	0	1	0.97	-3	50	48	-4
Vanadium	0	0.0	0	5	5.2	4	50	50	0
Zinc	0	0.0	0	5	5.8	16	500	510	2

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No.: _____ SDG No.: MYCPM3
 Instrument ID: P7 Start Date: 12/19/2024
 Analytical Method: ICP-MS Run Batch: LB134025
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	2500	2400	-4	5000	5000	0	10000	10000	0
Antimony	125	130	4	250	260	4	500	510	2
Arsenic	125	130	4	250	260	4	500	510	2
Barium	625	650	4	1250	1300	4	2500	2500	0
Beryllium	125	130	4	250	250	0	500	490	-2
Cadmium	125	130	4	250	260	4	500	510	2
Calcium	12500	13000	4	25000	25000	0	50000	51000	2
Chromium	125	130	4	250	250	0	500	510	2
Cobalt	125	130	4	250	260	4	500	510	2
Copper	1250	1300	4	2500	2600	4	5000	5100	2
Iron	6250	6800	9	12500	13000	4	25000	26000	4
Lead	625	630	1	1250	1300	4	2500	2500	0
Magnesium	12500	13000	4	25000	25000	0	50000	50000	0
Manganese	1250	1300	4	2500	2600	4	5000	5100	2
Nickel	125	130	4	250	250	0	500	500	0
Potassium	6250	6300	1	12500	13000	4	25000	25000	0
Selenium	125	130	4	250	260	4	500	510	2
Silver	125	140	12	250	260	4	500	520	4
Sodium	12500	13000	4	25000	26000	4	50000	51000	2
Thallium	125	130	4	250	250	0	500	500	0
Vanadium	125	130	4	250	250	0	500	510	2
Zinc	1250	1300	4	2500	2600	4	5000	5200	4

FORM 15-IN
INITIAL CALIBRATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No.: _____ SDG No.: MYCPM3
 Instrument ID: P7 Start Date: 12/19/2024
 Analytical Method: ICP-MS Run Batch: LB134025
 Concentration Units: ug/L

Analyte	True	Found	%D	True	Found	%D	True	Found	%D
Aluminum	20000	19000	-5	100000	100000	0			
Antimony	1000	990	-1						
Arsenic	1000	1000	0						
Barium	5000	5000	0						
Beryllium	1000	1000	0						
Cadmium	1000	990	-1						
Calcium	100000	100000	0	500000	500000	0			
Chromium	1000	990	-1						
Cobalt	1000	990	-1						
Copper	10000	9900	-1						
Iron	50000	51000	2	250000	250000	0			
Lead	5000	5000	0						
Magnesium	100000	97000	-3	500000	500000	0			
Manganese	10000	9900	-1						
Nickel	1000	1000	0						
Potassium	50000	49000	-2	250000	250000	0			
Selenium	1000	990	-1						
Silver	1000	990	-1						
Sodium	100000	99000	-1	500000	500000	0			
Thallium	1000	1000	0						
Vanadium	1000	1000	0						
Zinc	10000	9900	-1						