

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZG9Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134057 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID: ICV009				ID: CCV081				ID: CCV082		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum	2500	2400	96	1	400000	410000	102	0	400000	101	0
Antimony	1000	1000	102	0	5000	4900	98	0	4800	96	0
Arsenic	1000	1100	110	0	5000	5000	100	0	4800	96	0
Barium	520	520	101	0	10000	10000	103	0	10000	101	0
Beryllium	510	530	103	1	500	520	103	0	510	101	0
Cadmium	510	500	99	0	2500	2500	100	0	2400	97	0
Calcium	10000	10000	104	0	400000	410000	103	0	400000	101	0
Chromium	520	550	107	0	15500	16000	101	1	15000	99	0
Cobalt	520	500	97	0	2500	2500	100	0	2400	98	0
Copper	510	510	100	0	15000	15000	102	0	15000	100	0
Iron	10000	11000	105	0	400000	420000	104	0	410000	102	0
Lead	1000	980	98	1	25000	25000	100	0	24000	97	0
Magnesium	6000	6300	105	2	400000	410000	102	0	400000	101	0
Manganese	520	530	103	1	15000	15000	102	0	15000	100	0
Nickel	530	510	97	0	2500	2500	101	0	2500	98	0
Potassium	9900	9200	93	3	150000	150000	103	0	150000	101	0
Selenium	1000	950	95	1	5000	4900	98	0	4800	95	0
Silver	250	240	97	0	1250	1300	104	1	1300	103	1
Sodium	10000	9100	91	4	400000	410000	102	1	410000	102	0
Thallium	1000	1000	102	0	5000	4900	99	0	4800	95	0
Vanadium	500	510	103	2	2500	2600	103	0	2500	100	0
Zinc	1000	940	94	1	15000	15000	99	1	14000	94	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZG9Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134057 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV083				ID: CCV084		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	410000	103	0	400000	101	1
Antimony					5000	5000	100	0	4800	97	0
Arsenic					5000	5100	103	0	4900	98	0
Barium					10000	10000	103	1	10000	101	1
Beryllium					500	520	103	1	510	101	1
Cadmium					2500	2500	100	0	2500	99	0
Calcium					400000	410000	104	0	400000	101	1
Chromium					15500	16000	102	1	15000	99	0
Cobalt					2500	2500	100	0	2500	99	0
Copper					15000	15000	103	0	15000	100	1
Iron					400000	420000	104	1	410000	102	1
Lead					25000	25000	100	0	25000	98	0
Magnesium					400000	410000	103	0	400000	100	1
Manganese					15000	15000	102	0	15000	99	1
Nickel					2500	2500	101	0	2500	100	0
Potassium					150000	150000	103	0	150000	100	1
Selenium					5000	5000	99	0	4800	97	0
Silver					1250	1300	106	1	1300	103	0
Sodium					400000	410000	102	0	410000	101	2
Thallium					5000	4800	97	0	4800	96	0
Vanadium					2500	2600	103	0	2500	100	1
Zinc					15000	15000	102	0	15000	98	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZG9Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134057 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV085				ID: CCV086		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	100	1	400000	101	0
Antimony					5000	4800	95	0	4800	97	0
Arsenic					5000	4900	97	0	4800	96	0
Barium					10000	10000	101	1	10000	102	0
Beryllium					500	500	100	1	510	101	0
Cadmium					2500	2400	98	0	2500	99	0
Calcium					400000	400000	101	0	410000	101	0
Chromium					15500	15000	97	0	15000	99	0
Cobalt					2500	2400	97	0	2500	99	0
Copper					15000	15000	99	0	15000	100	0
Iron					400000	410000	101	1	410000	102	0
Lead					25000	24000	97	0	24000	98	0
Magnesium					400000	400000	100	1	400000	101	0
Manganese					15000	15000	99	1	15000	100	0
Nickel					2500	2500	99	0	2500	99	0
Potassium					150000	150000	100	0	150000	100	0
Selenium					5000	4800	96	0	4800	96	0
Silver					1250	1300	102	0	1300	103	0
Sodium					400000	400000	101	0	400000	101	0
Thallium					5000	4700	95	0	4700	95	0
Vanadium					2500	2500	100	0	2500	100	0
Zinc					15000	15000	99	1	15000	100	1

FORM 2 - IN

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011Lab Code: ACE Case No.: 51779 MA No. : SDG No.: MYCZG9Initial Calibration Verification Source : MP83508Continuing Calibration Verification Source : MP83511Run Batch: LB134057 Analytical Method: ICP-AESConcentration Units: µg/L

	Initial Calibration Verification				Continuing Calibration Verification						
	ID:				ID: CCV087				ID:		
Analyte	True	Found	%R	%RSD	True	Found	%R	%RSD	Found	%R	%RSD
Aluminum					400000	400000	100	1			
Antimony					5000	4800	95	0			
Arsenic					5000	4800	96	0			
Barium					10000	10000	101	1			
Beryllium					500	500	100	1			
Cadmium					2500	2400	98	0			
Calcium					400000	400000	100	0			
Chromium					15500	15000	98	0			
Cobalt					2500	2500	98	0			
Copper					15000	15000	99	1			
Iron					400000	400000	100	1			
Lead					25000	24000	97	0			
Magnesium					400000	400000	99	1			
Manganese					15000	15000	99	1			
Nickel					2500	2500	99	0			
Potassium					150000	150000	99	1			
Selenium					5000	4700	95	0			
Silver					1250	1300	102	0			
Sodium					400000	400000	99	1			
Thallium					5000	4700	94	1			
Vanadium					2500	2500	99	1			
Zinc					15000	15000	98	1			