

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51779MA No.: _____ SDG No.: MYCZH9Instrument ID: P5Start Date : 12/19/2024Analytical Method: ICP-AESRun Batch : LB134036

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999987	0.096440	-0.000666	WLR Blank	Inverse Conc
Antimony	0.997891	0.112706	0.000246	WLR Blank	Inverse Conc
Arsenic	0.999921	108.232040	0.236807	WLR Blank	Inverse Conc
Barium	0.999977	3.639615	0.020714	WLR Blank	Inverse Conc
Beryllium	0.999944	11.335668	0.000819	WLR Blank	Inverse Conc
Boron	0.999499	0.648136	0.005538	WLR Blank	Inverse Conc
Cadmium	0.999919	3.476188	0.000502	WLR Blank	Inverse Conc
Calcium	0.999979	0.045497	0.007346	WLR Blank	Inverse Conc
Chromium	0.999981	0.172803	0.000115	WLR Blank	Inverse Conc
Cobalt	0.999967	0.975747	-0.000273	WLR Blank	Inverse Conc
Copper	0.999971	0.183918	0.002590	WLR Blank	Inverse Conc
Iron	0.999781	0.021627	-0.000156	WLR Blank	Inverse Conc
Lead	0.999776	0.220115	0.000593	WLR Blank	Inverse Conc
Lithium	0.999985	1.092978	0.005832	WLR Blank	Inverse Conc
Magnesium	0.999943	0.008384	-0.001181	WLR Blank	Inverse Conc
Manganese	0.999925	0.497428	0.000149	WLR Blank	Inverse Conc
Molybdenum	0.998395	0.593859	0.000083	WLR Blank	Inverse Conc
Nickel	0.999917	0.698897	-0.000385	WLR Blank	Inverse Conc
Phosphorus	0.999403	0.077765	-0.003054	WLR Blank	Inverse Conc
Potassium	0.999818	0.025942	0.100824	WLR Blank	Inverse Conc
Selenium	0.998717	0.071602	0.000038	WLR Blank	Inverse Conc
Silicon	0.994523	1.456542	0.013213	WLR Blank	Inverse Conc
Silver	0.999848	0.236425	-0.001877	WLR Blank	Inverse Conc
Sodium	0.999953	0.006318	-0.021785	WLR Blank	Inverse Conc
Strontium	0.999978	8.495734	-0.004889	WLR Blank	Inverse Conc
Sulfur	0.999112	0.049186	0.000541	WLR Blank	Inverse Conc
Thallium	0.999975	0.066788	-0.000059	WLR Blank	Inverse Conc
Tin	0.999744	0.149349	0.001499	WLR Blank	Inverse Conc
Vanadium	0.999992	0.107632	0.000223	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51779 MA No.: SDG No.: MYCZH9
Instrument ID: P5 Start Date : 12/19/2024
Analytical Method: ICP-AES Run Batch : LB134036

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Zinc	0.998734	0.031660	-0.000558	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51779 MA No.: SDG No.: MYCZH9
Instrument ID: P5 Start Date : 12/19/2024
Analytical Method: ICP-AES Run Batch : LB134036

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------