

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No.: SDG No.: MYCZR9
 Instrument ID: P7 Date: 12/20/2024
 Run Batch: LB134058

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.1	0.7	5	0.41
Co - 59	59.0	0.8	5	0.99
In - 113	113.0	0.7	5	0.41
In - 115	115.0	0.8	5	0.65
Mg - 24	23.9	0.8	5	0.74
Mg - 25	24.9	0.7	5	0.29
Mg - 26	25.9	0.7	5	0.30
Pb - 206	206.0	0.7	5	0.52
Pb - 207	207.0	0.7	5	0.42
Pb - 208	208.0	0.8	5	0.69

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No.: SDG No.: MYCZR9
 Instrument ID: P7 Date: 12/20/2024
 Run Batch: LB134059

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.1	0.7	5	0.25
Co - 59	59.0	0.8	5	0.80
In - 113	113.0	0.7	5	0.46
In - 115	115.0	0.7	5	0.97
Mg - 24	23.9	0.8	5	0.44
Mg - 25	24.9	0.7	5	0.29
Mg - 26	25.9	0.7	5	0.15
Pb - 206	206.0	0.7	5	0.74
Pb - 207	206.9	0.7	5	0.52
Pb - 208	207.9	0.7	5	2.7

FORM 13-IN
ICP-MS TUNE

EPA SAMPLE NO.

TUNE

Lab Name: Alliance Technical Group, L Contract: 68HERH20D0011
 Lab Code: ACE Case No.: 51779 MA No.: SDG No.: MYCZR9
 Instrument ID: P7 Date: 12/23/2024
 Run Batch: LB134068

Element - Mass	Avg. Measured Mass (u)	Average Peak Width (u)	%Height	%RSD
Be - 9	9.1	0.7	5	0.56
Co - 59	59.0	0.8	5	0.86
In - 113	113.0	0.7	5	0.70
In - 115	115.0	0.8	5	1.1
Mg - 24	24.0	0.7	5	0.72
Mg - 25	25.0	0.8	5	0.49
Mg - 26	26.0	0.8	5	0.67
Pb - 206	206.1	0.8	5	0.41
Pb - 207	207.1	0.7	5	0.59
Pb - 208	208.1	0.8	5	0.64