

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51779 MA No. :                      SDG No.: MYCZR9  
 Analytical Method: ICP-MS ICSA Source: EPA-0803  
 Instrument ID: P7 ICSB Source: EPA-0803  
 Run Batch: LB134058  
 Concentration Units: ug/L

| Analyte   | True   |        | Found  |     |        |     |
|-----------|--------|--------|--------|-----|--------|-----|
|           | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Aluminum  | 100000 | 100000 | 93000  | 93  | 93000  | 93  |
| Antimony  | 1.5    | 22     | 1.1    | 76  | 21     | 96  |
| Arsenic   | 0.1    | 19     | 0.36   | 360 | 21     | 112 |
| Barium    | 1.2    | 22     | 1.4    | 117 | 21     | 97  |
| Beryllium | 0      | 19     | 0.29   |     | 22     | 113 |
| Cadmium   | 0.7    | 20     | 0.35   | 50  | 20     | 100 |
| Calcium   | 100000 | 100000 | 100000 | 100 | 100000 | 102 |
| Chromium  | 21     | 40     | 21     | 98  | 41     | 102 |
| Cobalt    | 1      | 20     | 1.3    | 128 | 22     | 108 |
| Copper    | 8      | 25     | 7.2    | 91  | 26     | 105 |
| Iron      | 100000 | 100000 | 110000 | 105 | 110000 | 105 |
| Lead      | 4      | 25     | 4.8    | 119 | 24     | 95  |
| Magnesium | 100000 | 100000 | 99000  | 99  | 98000  | 98  |
| Manganese | 7      | 27     | 8.1    | 116 | 28     | 105 |
| Nickel    | 6      | 24     | 5.4    | 90  | 26     | 106 |
| Potassium | 100000 | 100000 | 99000  | 99  | 99000  | 99  |
| Selenium  | 0.3    | 19     | -0.12  | 0   | 19     | 100 |
| Silver    | 0      | 18     | 0.04   |     | 19     | 107 |
| Sodium    | 100000 | 100000 | 100000 | 100 | 100000 | 100 |
| Thallium  | 0      | 21     | 0.14   |     | 20     | 95  |
| Vanadium  | 0.5    | 19     | 0.2    | 40  | 20     | 107 |
| Zinc      | 11     | 29     | 12     | 105 | 32     | 111 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51779 MA No. :                      SDG No.: MYCZR9  
 Analytical Method: ICP-MS ICSA Source: EPA-0803  
 Instrument ID: P7 ICSB Source: EPA-0803  
 Run Batch: LB134059  
 Concentration Units: ug/L

| Analyte   | True   |        | Found  |     |        |     |
|-----------|--------|--------|--------|-----|--------|-----|
|           | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Aluminum  | 100000 | 100000 | 92000  | 92  | 92000  | 92  |
| Antimony  | 1.5    | 22     | 1.1    | 75  | 21     | 96  |
| Arsenic   | 0.1    | 19     | 0.34   | 340 | 21     | 111 |
| Barium    | 1.2    | 22     | 1.4    | 117 | 22     | 98  |
| Beryllium | 0      | 19     | 0.3    |     | 22     | 114 |
| Cadmium   | 0.7    | 20     | 0.36   | 51  | 20     | 100 |
| Calcium   | 100000 | 100000 | 100000 | 101 | 100000 | 101 |
| Chromium  | 21     | 40     | 20     | 96  | 41     | 102 |
| Cobalt    | 1      | 20     | 1.3    | 128 | 22     | 108 |
| Copper    | 8      | 25     | 7.2    | 90  | 26     | 105 |
| Iron      | 100000 | 100000 | 110000 | 105 | 110000 | 105 |
| Lead      | 4      | 25     | 4.5    | 113 | 24     | 95  |
| Magnesium | 100000 | 100000 | 97000  | 97  | 97000  | 97  |
| Manganese | 7      | 27     | 8.2    | 117 | 28     | 105 |
| Nickel    | 6      | 24     | 5.5    | 91  | 26     | 107 |
| Potassium | 100000 | 100000 | 98000  | 97  | 98000  | 98  |
| Selenium  | 0.3    | 19     | -0.12  | 0   | 21     | 112 |
| Silver    | 0      | 18     | 0.05   |     | 19     | 106 |
| Sodium    | 100000 | 100000 | 99000  | 99  | 99000  | 99  |
| Thallium  | 0      | 21     | 0.11   |     | 20     | 95  |
| Vanadium  | 0.5    | 19     | 0.21   | 42  | 21     | 108 |
| Zinc      | 11     | 29     | 11     | 100 | 31     | 108 |

FORM 4 - IN  
ICP INTERFERENCE CHECK SAMPLE

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011  
 Lab Code: ACE Case No.: 51779 MA No. :                      SDG No.: MYCZR9  
 Analytical Method: ICP-MS ICSA Source: EPA-0803  
 Instrument ID: P7 ICSB Source: EPA-0803  
 Run Batch: LB134068  
 Concentration Units: ug/L

| Analyte   | True   |        | Found  |     |        |     |
|-----------|--------|--------|--------|-----|--------|-----|
|           | ICSA   | ICSAB  | ICSA   | %R  | ICSAB  | %R  |
| Aluminum  | 100000 | 100000 | 93000  | 93  | 92000  | 92  |
| Antimony  | 1.5    | 22     | 1.1    | 74  | 21     | 94  |
| Arsenic   | 0.1    | 19     | 0.35   | 350 | 21     | 112 |
| Barium    | 1.2    | 22     | 1.4    | 118 | 21     | 96  |
| Beryllium | 0      | 19     | 0.3    |     | 21     | 111 |
| Cadmium   | 0.7    | 20     | 0.22   | 31  | 20     | 97  |
| Calcium   | 100000 | 100000 | 100000 | 103 | 100000 | 101 |
| Chromium  | 21     | 40     | 21     | 99  | 41     | 102 |
| Cobalt    | 1      | 20     | 1.3    | 126 | 21     | 107 |
| Copper    | 8      | 25     | 7.2    | 90  | 26     | 104 |
| Iron      | 100000 | 100000 | 110000 | 105 | 100000 | 102 |
| Lead      | 4      | 25     | 4.6    | 114 | 23     | 93  |
| Magnesium | 100000 | 100000 | 95000  | 95  | 93000  | 93  |
| Manganese | 7      | 27     | 8.4    | 119 | 28     | 104 |
| Nickel    | 6      | 24     | 5.5    | 91  | 25     | 105 |
| Potassium | 100000 | 100000 | 100000 | 100 | 99000  | 98  |
| Selenium  | 0.3    | 19     | -0.18  | 0   | 20     | 107 |
| Silver    | 0      | 18     | 0.04   |     | 19     | 105 |
| Sodium    | 100000 | 100000 | 100000 | 100 | 99000  | 99  |
| Thallium  | 0      | 21     | 0.1    |     | 19     | 91  |
| Vanadium  | 0.5    | 19     | 0.17   | 34  | 20     | 108 |
| Zinc      | 11     | 29     | 11     | 104 | 32     | 109 |