

Initial and Continuing Calibration Verification

Client: Ardmore Chemical

SDG No.: P5192

Project: PVSC Monthly 2024

RunNo.: LB133924

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: Cyanide	ICV1	mg/L	0.096	0.099	97	85-115	12/12/2024
Sample ID: Cyanide	CCV1	mg/L	0.24	0.25	96	90-110	12/12/2024
Sample ID: Cyanide	CCV2	mg/L	0.25	0.25	100	90-110	12/12/2024