



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: Ardmore Chemical **SDG No.:** P5192
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** P5192 **SAS No.:** P5192
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.038	0.05	76	40 - 160	CV	12/10/2024	10:39	LB133852
CRI01	Cadmium	5.88	6.0	98	40 - 160	P	12/20/2024	16:26	LB134054
	Copper	21.8	20.0	109	40 - 160	P	12/20/2024	16:26	LB134054
	Lead	11.1	12.0	93	40 - 160	P	12/20/2024	16:26	LB134054
	Nickel	39.0	40.0	98	40 - 160	P	12/20/2024	16:26	LB134054
	Zinc	38.8	40.0	97	40 - 160	P	12/20/2024	16:26	LB134054