



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	Ardmore Chemical				SDG No.:	P5192			
Contract:	ARDM01	Lab Code:	CHEM		Case No.:	P5192	SAS No.:	P5192	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB14	Mercury	0.20	+/-0.20	U	0.20	CV	12/10/2024	10:17	LB133852

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: P5192

Contract: ARDM01

Lab Code: CHEM

Case No.: P5192

SAS No.: P5192

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB10	Mercury	0.20	+/-0.20	U	0.20	CV	12/10/2024	10:34	LB133852
CCB11	Mercury	0.20	+/-0.20	U	0.20	CV	12/10/2024	11:31	LB133852
CCB12	Mercury	0.20	+/-0.20	U	0.20	CV	12/10/2024	11:49	LB133852

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: P5192

Contract: ARDM01

Lab Code: CHEM

Case No.: P5192

SAS No.: P5192

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	16:21	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	16:21	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	16:21	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	16:21	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	16:21	LB134054

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>Ardmore Chemical</u>		SDG No.: <u>P5192</u>							
Contract: <u>ARDM01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>P5192</u>	SAS No.: <u>P5192</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	16:51	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	16:51	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	16:51	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	16:51	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	16:51	LB134054
CCB02	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	17:34	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	17:34	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	17:34	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	17:34	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	17:34	LB134054
CCB03	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	18:16	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	18:16	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	18:16	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	18:16	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	18:16	LB134054
CCB04	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	19:11	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	19:11	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	19:11	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	19:11	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	19:11	LB134054
CCB05	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	20:01	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	20:01	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	20:01	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	20:01	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	20:01	LB134054
CCB06	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	20:51	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	20:51	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	20:51	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	20:51	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	20:51	LB134054
CCB07	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	22:04	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	22:04	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	22:04	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	22:04	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	22:04	LB134054
CCB08	Cadmium	6.00	+/-6.00	U	6.00	P	12/20/2024	23:12	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/20/2024	23:12	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/20/2024	23:12	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/20/2024	23:12	LB134054

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: Ardmore Chemical

SDG No.: P5192

Contract: ARDM01

Lab Code: CHEM

Case No.: P5192

SAS No.: P5192

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB08	Zinc	40.0	+/-40.0	U	40.0	P	12/20/2024	23:12	LB134054
CCB09	Cadmium	6.00	+/-6.00	U	6.00	P	12/21/2024	00:05	LB134054
	Copper	20.0	+/-20.0	U	20.0	P	12/21/2024	00:05	LB134054
	Lead	12.0	+/-12.0	U	12.0	P	12/21/2024	00:05	LB134054
	Nickel	40.0	+/-40.0	U	40.0	P	12/21/2024	00:05	LB134054
	Zinc	40.0	+/-40.0	U	40.0	P	12/21/2024	00:05	LB134054