

Metals
- 4 -
INTERFERENCE CHECK SAMPLE

Client: Ardmore Chemical **SDG No.:** P5192
Contract: ARDM01 **Lab Code:** CHEM **Case No.:** P5192 **SAS No.:** P5192
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Cadmium	-0.99	1.0	99	-5	7	12/20/2024	16:30	LB134054
	Copper	16.4	2.0	820	-18	22	12/20/2024	16:30	LB134054
	Lead	6.97			-12	12	12/20/2024	16:30	LB134054
	Nickel	2.13	2.0	106	-38	42	12/20/2024	16:30	LB134054
	Zinc	0.10			-40	40	12/20/2024	16:30	LB134054
ICSAB01	Cadmium	1040	972	107	826	1120	12/20/2024	16:34	LB134054
	Copper	520	511	102	434	588	12/20/2024	16:34	LB134054
	Lead	56.2	49.0	115	37	61	12/20/2024	16:34	LB134054
	Nickel	1040	954	109	810	1100	12/20/2024	16:34	LB134054
	Zinc	1080	952	113	809	1095	12/20/2024	16:34	LB134054
ICSA	Cadmium	-2.22	1.0	222	-5	7	12/20/2024	16:38	LB134054
	Copper	18.5	2.0	925	-18	22	12/20/2024	16:38	LB134054
	Lead	-1.48			-12	12	12/20/2024	16:38	LB134054
	Nickel	-0.44	2.0	22	-38	42	12/20/2024	16:38	LB134054
	Zinc	-51.8			-40	40	12/20/2024	16:38	LB134054
ICSAB	Cadmium	1050	972	108	826	1120	12/20/2024	16:43	LB134054
	Copper	561	511	110	434	588	12/20/2024	16:43	LB134054
	Lead	59.9	49.0	122	37	61	12/20/2024	16:43	LB134054
	Nickel	1060	954	111	810	1100	12/20/2024	16:43	LB134054
	Zinc	1040	952	109	809	1095	12/20/2024	16:43	LB134054