

Surrogate Summary

SW-846

SDG No.: **P5192**

Client: **Ardmore Chemical**

Analytical Method: **625.1**

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P5192-02	EFF-WASTE WATER	2-Fluorophenol	100	54.2	54	*	60	140
		Phenol-d6	100	34.5	35	*	60	140
		Nitrobenzene-d5	100	116	116		60	140
		2-Fluorobiphenyl	100	91.6	92		60	140
		2,4,6-Tribromophenol	100	86.2	86		60	140
		Terphenyl-d14	100	89.5	89		60	140
P5192-02RE	EFF-WASTE WATERRE	2-Fluorophenol	100	48.6	49	*	60	140
		Phenol-d6	100	30.4	30	*	60	140
		Nitrobenzene-d5	100	106	106		60	140
		2-Fluorobiphenyl	100	90.9	91		60	140
		2,4,6-Tribromophenol	100	85.7	86		60	140
		Terphenyl-d14	100	85.3	85		60	140
PB165479BL	PB165479BL	2-Fluorophenol	100	85.2	85		60	140
		Phenol-d6	100	86.3	86		60	140
		Nitrobenzene-d5	100	93.7	94		60	140
		2-Fluorobiphenyl	100	84.5	84		60	140
		2,4,6-Tribromophenol	100	76.4	76		60	140
		Terphenyl-d14	100	87.6	88		60	140
PB165479BS	PB165479BS	2-Fluorophenol	100	96.3	96		60	140
		Phenol-d6	100	92.2	92		60	140
		Nitrobenzene-d5	100	96.9	97		60	140
		2-Fluorobiphenyl	100	94.2	94		60	140
		2,4,6-Tribromophenol	100	94.2	94		60	140
		Terphenyl-d14	100	94.4	94		60	140
PB165479BSD	PB165479BSD	2-Fluorophenol	100	99.2	99		60	140
		Phenol-d6	100	90.7	91		60	140
		Nitrobenzene-d5	100	93.6	94		60	140
		2-Fluorobiphenyl	100	89.3	89		60	140
		2,4,6-Tribromophenol	100	90.2	90		60	140
		Terphenyl-d14	100	87.5	87		60	140