

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARDM01

Lab Code: CHEM Case No.: P5192 SAS No.: P5192 SDG No.: P5192

Instrument ID: MSVOA_N Calibration Date(s): 12/06/2024 12/06/2024

Heated Purge: (Y/N) N Calibration Time(s): 10:02 11:37

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VN085117.D		RRF020 = VN085118.D		RRF050 = VN085119.D			
		RRF100 = VN085120.D		RRF150 = VN085121.D		RRF =			
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	RRF	% RSD	
Chloromethane	2.058	2.093	2.013	2.202	2.256		2.124	4.8	
Vinyl Chloride	2.171	2.147	2.026	2.258	2.263		2.173	4.5	
Bromomethane	1.578	1.334	1.261	1.364	1.361		1.380	8.6	
Chloroethane	1.331	1.302	1.259	1.382	1.401		1.335	4.4	
Trichlorofluoromethane	3.495	3.461	3.321	3.725	3.781		3.557	5.4	
1,1-Dichloroethene	1.669	1.752	1.743	1.978	1.993		1.827	8.1	
Acrolein	0.309	0.255	0.299	0.338	0.365		0.313	13.2	
Acrylonitrile	0.809	0.837	0.858	0.977	0.994		0.895	9.4	
Methylene Chloride	2.198	2.012	1.927	2.127	2.100		2.073	5.1	
trans-1,2-Dichloroethene	1.876	1.799	1.782	2.012	2.027		1.899	6.1	
1,1-Dichloroethane	3.621	3.416	3.269	3.705	3.693		3.541	5.4	
Carbon Tetrachloride	0.619	0.587	0.566	0.591	0.593		0.591	3.2	
Chloroform	3.779	3.635	3.508	3.813	3.814		3.710	3.6	
1,1,1-Trichloroethane	0.683	0.665	0.619	0.657	0.652		0.655	3.6	
Benzene	1.581	1.501	1.456	1.543	1.523		1.521	3.1	
1,2-Dichloroethane	0.564	0.505	0.491	0.518	0.510		0.518	5.4	
Trichloroethene	0.367	0.356	0.344	0.373	0.371		0.362	3.4	
1,2-Dichloropropane	0.388	0.370	0.351	0.366	0.368		0.368	3.7	
Bromodichloromethane	0.596	0.554	0.541	0.565	0.568		0.565	3.6	
Toluene	1.780	1.713	1.680	1.799	1.777		1.750	2.9	
t-1,3-Dichloropropene	0.518	0.517	0.541	0.592	0.603		0.554	7.4	
cis-1,3-Dichloropropene	0.581	0.590	0.583	0.629	0.636		0.604	4.4	
1,1,2-Trichloroethane	0.387	0.355	0.336	0.353	0.349		0.356	5.2	
2-Chloroethyl vinyl ether	0.200	0.247	0.258	0.284	0.269		0.252	12.7	
Dibromochloromethane	0.455	0.426	0.413	0.438	0.438		0.434	3.7	
Tetrachloroethene	0.407	0.358	0.342	0.359	0.351		0.363	7	
Chlorobenzene	1.074	1.056	1.037	1.121	1.113		1.080	3.3	
Ethyl Benzene	1.626	1.739	1.845	2.032	2.041		1.857	9.8	
m/p-Xylenes	0.601	0.713	0.706	0.775	0.775		0.714	10	
o-Xylene	0.561	0.648	0.681	0.734	0.729		0.671	10.5	

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: ARDM01
 Lab Code: CHEM Case No.: P5192 SAS No.: P5192 SDG No.: P5192
 Instrument ID: MSVOA_N Calibration Date(s): 12/06/2024 12/06/2024
 Heated Purge: (Y/N) N Calibration Time(s): 10:02 11:37
 GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VN085117.D		RRF020 = VN085118.D		RRF050 = VN085119.D	
		RRF100 = VN085120.D		RRF150 = VN085121.D		RRF =	
COMPOUND	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Bromoform	0.294	0.275	0.282	0.301	0.302	0.291	4.1
1,1,2,2-Tetrachloroethane	0.551	0.532	0.503	0.526	0.521	0.526	3.3
1,3-Dichlorobenzene	0.779	0.765	0.772	0.842	0.816	0.795	4.1
1,4-Dichlorobenzene	0.725	0.742	0.759	0.829	0.818	0.775	6
1,2-Dichlorobenzene	0.743	0.721	0.734	0.787	0.767	0.750	3.5
1,2-Dichloroethane-d4	2.462	2.393	2.263	2.370	2.414	2.380	3.1
Toluene-d8	1.462	1.443	1.392	1.372	1.367	1.407	3.1
4-Bromofluorobenzene	0.479	0.495	0.500	0.509	0.499	0.496	2.2

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.