

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112124\
 Data File : PR069313.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 21 Nov 2024 21:07
 Operator : AJ\MA
 Sample : AR1248ICC100
 Misc :
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 ECD_R
 ClientSampleId :
 AR12481214

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/22/2024
 Supervised By :Ankita Jodhani 11/22/2024

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Nov 22 02:19:00 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Fri Nov 22 02:18:33 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 1 µl
 Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
 Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds						
1) SA Tetrachlo...	3.637	2.969	5471543	37292739	5.366	5.765
2) SA Decachlor...	9.531	8.253	6003450	41773867	11.016	10.936
Target Compounds						
21) L5 AR-1248-1	4.857	4.090	2309753	14163188	118.739m	111.502
22) L5 AR-1248-2	5.150	4.340	3215744	20172458	118.212	116.450
23) L5 AR-1248-3	5.361	4.382	3617293	20282415	113.586	116.333
24) L5 AR-1248-4	5.794	4.558	3928504	23667307	112.859	116.087
25) L5 AR-1248-5	5.834	4.973	3800748	22574413	109.089	116.358

(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_R\Data\PR112124\
Data File : PR069313.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 21 Nov 2024 21:07
Operator : AJ\MA
Sample : AR1248ICC100
Misc :
ALS Vial : 23 Sample Multiplier: 1

Instrument :
ECD_R
ClientSampleId :
AR12481214

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/22/2024
Supervised By :Ankita Jodhani 11/22/2024

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Nov 22 02:19:00 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_R\Method\PR112124CLP.M
Quant Title : GC EXTRACTABLES
QLast Update : Fri Nov 22 02:18:33 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 1 µl
Signal #1 Phase : ZB MR1 Signal #2 Phase: ZB MR2
Signal #1 Info : 30Mx0.32mmx 0.5µm Signal #2 Info : 30M x 0.32mm x 0.25µm

