

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5193 SAS No.: P5193 SDG NO.: P5193

EPA Sample No.: SSTD020741 Date Analyzed: 12/10/2024

Lab File ID: BP023355.D Time Analyzed: 21:17

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	98691	7.546	388630	10.29	301279	14.16
UPPER LIMIT	197382	8.046	777260	10.787	602558	14.657
LOWER LIMIT	49345.5	7.046	194315	9.787	150640	13.657
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	102513	7.54	372899	10.29	257805	14.16

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.