

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P5193 SAS No.: P5193 SDG NO.: P5193

EPA Sample No.: SSTD020741 Date Analyzed: 12/10/2024

Lab File ID: BP023355.D Time Analyzed: 21:17

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	727548	16.963	786712	21.38	876155	24.551
UPPER LIMIT	1455100	17.463	1573420	21.88	1752310	25.051
LOWER LIMIT	363774	16.463	393356	20.88	438078	24.051
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	572704	16.96	614013	21.39	699851	24.57

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.