

Surrogate Summary

SDG No.: **P5193**

Client: **Chemtech Consulting Group**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5193-01	Storage-Blank-SOIL-REF-6	1,2-Dichloroethane-d4	50	49.8	100	74	125
		Dibromofluoromethane	50	46.5	93	75	124
		Toluene-d8	50	49.5	99	86	113
		4-Bromofluorobenzene	50	48.5	97	77	121
P5193-02	Storage-Blank-WATER-REF-5	1,2-Dichloroethane-d4	50	50.0	100	74	125
		Dibromofluoromethane	50	46.6	93	75	124
		Toluene-d8	50	49.3	99	86	113
		4-Bromofluorobenzene	50	47.7	95	77	121
P5193-03	Storage-Blank-WATER-REF-4	1,2-Dichloroethane-d4	50	50.7	101	74	125
		Dibromofluoromethane	50	45.8	92	75	124
		Toluene-d8	50	48.4	97	86	113
		4-Bromofluorobenzene	50	48.2	96	77	121
P5193-04	Storage-Blank-SAMPLE-MANAGEMENT	1,2-Dichloroethane-d4	50	49.0	98	74	125
		Dibromofluoromethane	50	46.2	92	75	124
		Toluene-d8	50	49.9	100	86	113
		4-Bromofluorobenzene	50	50.3	101	77	121
VX1210WBL01	VX1210WBL01	1,2-Dichloroethane-d4	50	49.7	99	74	125
		Dibromofluoromethane	50	46.4	93	75	124
		Toluene-d8	50	49.1	98	86	113
		4-Bromofluorobenzene	50	47.5	95	77	121
VX1210WBS01	VX1210WBS01	1,2-Dichloroethane-d4	50	52.6	105	74	125
		Dibromofluoromethane	50	50.9	102	75	124
		Toluene-d8	50	49.6	99	86	113
		4-Bromofluorobenzene	50	51.7	103	77	121
VX1210WBSD0	VX1210WBSD01	1,2-Dichloroethane-d4	50	51.3	103	74	125
		Dibromofluoromethane	50	51.1	102	75	124
		Toluene-d8	50	49.7	99	86	113
		4-Bromofluorobenzene	50	53.8	108	77	121