

DIESEL RANGE ORGANICS INITIAL CALIBRATION SUMMARY

Lab Name: Chemtech Contract: JPCL01
 ProjectID: 1454 to 1460 Haddon Avenue, Camden, NJ
 Lab Code: CHEM Case No.: P5194 SAS No.: P5194 SDG No.: P5194

Calibration Sequence : FG112024		Test : Diesel Range Organics	
Concentration (PPM)	Area Count	Reference Factor	File ID
1000	122483220	122483	FG014888.D
500	61587639	123175	FG014889.D
200	25738367	128692	FG014890.D
100	14303309	143033	FG014891.D
50	6841225	136824	FG014892.D
AVG RF : 130841		% RSD : 6.811	AVG RT : 15.0408