

DIESEL RANGE ORGANICS CONTINUING CALIBRATION SUMMARY

50 PPM TRPH STD

Lab Name: Chemtech Contract: JPCL01
ProjectID: 1454 to 1460 Haddon Avenue, Camden, NJ
Lab Code: CHEM Case No.: P5194 SAS No.: P5194 SDG No.: P5194
DataFile: FG014970.D Analyst Name: YP\AJ Analyst Date: 12-11-2024

Conc. (PPM)	Area Count	RF	Average RF	%D
500	65892601	131785	130841	0.721