

Continuing Calibration Report for SequenceID : FC121124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067983.D

Aliphatic C9-C12	8663037.000	60.000	3.151	6.428	144383.950	158302.459	8.792
Aliphatic C12-C16	6011580.000	40.000	6.429	9.820	150289.500	160275.290	6.230
Aliphatic C16-C21	8662899.000	60.000	9.821	13.180	144381.650	152362.625	5.238
Aliphatic C21-C28	10464742.000	80.000	13.181	16.836	130809.275	136384.122	4.088
Aliphatic C28-C40	12972101.000	120.000	16.837	21.674	108100.842	105964.639	-2.016
Aliphatic EPH	46774359.000	360.000	3.151	21.674	129928.775	135215.008	3.910