

Continuing Calibration Report for SequenceID : FC121124AL

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FC067993.D

Aliphatic C9-C12	9185556.000	60.000	3.151	6.428	153092.600	158302.459	3.291
Aliphatic C12-C16	6207242.000	40.000	6.429	9.820	155181.050	160275.290	3.178
Aliphatic C16-C21	8760777.000	60.000	9.821	13.180	146012.950	152362.625	4.167
Aliphatic C21-C28	10604423.000	80.000	13.181	16.836	132555.288	136384.122	2.807
Aliphatic C28-C40	13092661.000	120.000	16.837	21.674	109105.508	105964.639	-2.964
Aliphatic EPH	47850659.000	360.000	3.151	21.674	132918.497	135215.008	1.698