

Continuing Calibration Report for SequenceID : FD121124AR

Parameter	AreaCount	Conc.	RT_Min	RT_Max	Response Factor	AVGRF	%DEV
-----------	-----------	-------	--------	--------	-----------------	-------	------

File ID : FD048872.D

Aromatic C10-C12	7405186.000	40.000	4.086	5.804	185129.650	190209.437	2.671
Aromatic C12-C16	11067363.000	60.000	5.805	8.409	184456.050	185883.358	0.768
Aromatic C16-C21	13208158.000	80.000	8.410	12.673	165101.975	169519.943	2.606
Aromatic C21-C36	24543906.000	180.000	12.674	18.083	136355.033	141588.666	3.696
Aromatic EPH	56224613.000	360.000	4.086	18.083	156179.481	160580.373	2.741