

Continuing Calibration Report for SequenceID : FD121124AR

| Parameter | AreaCount | Conc. | RT_Min | RT_Max | Response Factor | AVGRF | %DEV |
|-----------|-----------|-------|--------|--------|-----------------|-------|------|
|-----------|-----------|-------|--------|--------|-----------------|-------|------|

File ID : FD048882.D

|                  |              |         |        |        |            |            |        |
|------------------|--------------|---------|--------|--------|------------|------------|--------|
| Aromatic C10-C12 | 6799290.000  | 40.000  | 4.086  | 5.804  | 169982.250 | 190209.437 | 10.634 |
| Aromatic C12-C16 | 10008221.000 | 60.000  | 5.805  | 8.409  | 166803.683 | 185883.358 | 10.264 |
| Aromatic C16-C21 | 11888636.000 | 80.000  | 8.410  | 12.673 | 148607.950 | 169519.943 | 12.336 |
| Aromatic C21-C36 | 23479063.000 | 180.000 | 12.674 | 18.083 | 130439.239 | 141588.666 | 7.875  |
| Aromatic EPH     | 52175210.000 | 360.000 | 4.086  | 18.083 | 144931.139 | 160580.373 | 9.745  |