

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** JPCL Engineering
Lab Code: CHEM **Cas No:** P5194 **SAS No :** P5194 **SDG No:** P5194
Sample No : P5194-01MS **Datafile:** FC067989.D
Client ID : UST-1MS

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aliphatic C9-C12	11.3	0	4.91	43		(40-140)
Aliphatic C12-C16	7.6	0.37	4.59	56		(40-140)
Aliphatic C16-C21	11.3	8.48	11.8	29	*	(40-140)
Aliphatic C21-C28	15.1	3.37	13.0	64		(40-140)
Aliphatic C28-C40	22.7	17.2	53.0	157	*	(40-140)