

SOLID EPH MATRIX SPIKE/MATRIX SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** JPCL Engineering
Lab Code: CHEM **Cas No:** P5194 **SAS No :** P5194 **SDG No:** P5194
Sample No : P5194-01MS **Datafile:** FD048878.D
Client ID : UST-1MS

COMPOUND	SPIKE ADDED mg/kg	SAMPLE CONCENTRATION mg/kg	MS/MSD CONCENTRATION mg/kg	% REC	Qual	QC LIMITS
Aromatic C10-C12	7.6	0	4.91	65		(40-140)
Aromatic C12-C16	11.3	0	11.7	103		(40-140)
Aromatic C16-C21	18.9	1.24	15.3	74		(40-140)
Aromatic C21-C36	30.2	4.04	23.6	65		(40-140)