

SOLID EPH LABORATORY CONTROL SPIKE/LABORATORY CONTROL SPIKE DUPLICATE RECOVERY

Lab Name: Chemtech **Client:** JPCL Engineering
Lab Code: CHEM **Cas No:** P5194 **SAS No :** P5194 **SDG No:** P5194
Sample No : PB165557BSD **Datafile:** FD048875.D
Client ID : PB165557BSD

COMPOUND	SPIKE ADDED mg/kg	LCS/LCSD CONCENTRATION mg/kg	% REC	Qual	RPD QC LIMITS	QC Limit Of RPD
Aromatic C10-C12	6.7	4.14	62		2.8 (40-140)	50
Aromatic C12-C16	10.0	7.86	79		3.3 (40-140)	50
Aromatic C16-C21	16.7	16.5	99		3 (40-140)	50
Aromatic C21-C36	26.7	26.3	99		3.5 (40-140)	50