

Surrogate Summary

SDG No.: **P5194**

Client: **JPCL Engineering**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P5194-03	WP-1	1,2-Dichloroethane-d4	50	63.3	127	50	163
		Dibromofluoromethane	50	52.8	106	54	147
		Toluene-d8	50	50.8	102	58	134
		4-Bromofluorobenzene	50	46.1	92	29	146
P5194-04	WP-2	1,2-Dichloroethane-d4	50	57.3	115	50	163
		Dibromofluoromethane	50	50.7	101	54	147
		Toluene-d8	50	50.8	102	58	134
		4-Bromofluorobenzene	50	45.1	90	29	146
P5194-05	WP-3	1,2-Dichloroethane-d4	50	59.5	119	50	163
		Dibromofluoromethane	50	51.5	103	54	147
		Toluene-d8	50	50.7	101	58	134
		4-Bromofluorobenzene	50	45.2	90	29	146
P5194-06	WP-4	1,2-Dichloroethane-d4	50	55.8	112	50	163
		Dibromofluoromethane	50	50.8	102	54	147
		Toluene-d8	50	49.8	100	58	134
		4-Bromofluorobenzene	50	43.8	88	29	146
VY1210SBL01	VY1210SBL01	1,2-Dichloroethane-d4	50	51.8	104	50	163
		Dibromofluoromethane	50	48.7	97	54	147
		Toluene-d8	50	49.9	100	58	134
		4-Bromofluorobenzene	50	42.9	86	29	146
VY1210SBS01	VY1210SBS01	1,2-Dichloroethane-d4	50	54.8	110	50	163
		Dibromofluoromethane	50	54.2	108	54	147
		Toluene-d8	50	52.5	105	58	134
		4-Bromofluorobenzene	50	55.0	110	29	146
VY1210SBSD01	VY1210SBSD01	1,2-Dichloroethane-d4	50	50.2	100	50	163
		Dibromofluoromethane	50	49.1	98	54	147
		Toluene-d8	50	47.8	96	58	134
		4-Bromofluorobenzene	50	48.3	97	29	146