

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JPCL01
 Lab Code: CHEM Case No.: P5194 SAS No.: P5194 SDG NO.: P5194
 Lab File ID: VY020560.D Date Analyzed: 12/10/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:42
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	176773	7.71	282168	8.62	241362	11.41
UPPER LIMIT	353546	8.207	564336	9.116	482724	11.914
LOWER LIMIT	88386.5	7.207	141084	8.116	120681	10.914
EPA SAMPLE NO.						
WP-1	97195	7.71	167005	8.62	147227	11.41
WP-2	154588	7.71	267938	8.61	230686	11.41
WP-3	144830	7.71	254650	8.61	223520	11.41
WP-4	151423	7.71	263038	8.61	223099	11.41
VY1210SBL01	179482	7.71	302822	8.62	250929	11.41
VY1210SBS01	155321	7.71	247103	8.61	209823	11.41
VY1210SBSD01	165679	7.71	267297	8.61	223946	11.41

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.