

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JPCL01
 Lab Code: CHEM Case No.: P5194 SAS No.: P5194 SDG NO.: P5194
 Lab File ID: VY020560.D Date Analyzed: 12/10/2024
 Instrument ID: MSVOA_Y Time Analyzed: 10:42
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	116370	13.347				
UPPER LIMIT	232740	13.847				
LOWER LIMIT	58185	12.847				
EPA SAMPLE NO.						
WP-1	59459	13.35				
WP-2	95035	13.35				
WP-3	90631	13.35				
WP-4	90370	13.35				
VY1210SBL01	97035	13.35				
VY1210SBS01	107923	13.35				
VY1210SBSD01	111842	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.