

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134256

Review By	Sarabjit Jaswal	Review On	1/13/2025 6:24:27 PM
Supervise By	Mohan Bera	Supervise On	1/13/2025 6:25:37 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501		
ICV Standard	MP83508		
CCV Standard	MP83511		
ICSA Standard	MP83509,MP83510		
CRI Standard			
LCS Standard	MP83506		
Chk Standard	MP83513,MP83514		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/13/25 15:23		Kareem	OK
2	S1	S01	CAL2	01/13/25 15:27		Kareem	OK
3	S2	S02	CAL3	01/13/25 15:31		Kareem	OK
4	S3	S03	CAL4	01/13/25 15:35		Kareem	OK
5	S4	S04	CAL5	01/13/25 15:39		Kareem	OK
6	S5	S05	CAL6	01/13/25 15:42		Kareem	OK
7	S6	S06	CAL7	01/13/25 15:47		Kareem	OK
8	ICV058	ICV058	ICV	01/13/25 16:02	Instrument paused for calibration review	Kareem	OK
9	ICB058	ICB058	ICB	01/13/25 16:06		Kareem	OK
10	ICSA058	ICSA058	ICSA	01/13/25 16:10		Kareem	OK
11	ICSAB058	ICSAB058	ICSAB	01/13/25 16:30	Instrument paused for review	Kareem	OK
12	CCV147	CCV147	CCV	01/13/25 16:41	Instrument paused for review	Kareem	OK
13	CCB147	CCB147	CCB	01/13/25 16:44		Kareem	OK
14	PB165922BL	PBS922	MB	01/13/25 16:53		Kareem	OK
15	PB165910BL	PBS910	MB	01/13/25 17:01		Kareem	OK
16	PB165910BS	LCS910	LCS	01/13/25 17:05		Kareem	OK
17	P5207-11	MBHN37	SAM	01/13/25 17:09		Kareem	OK
18	P5198-16	MBHMW3	SAM	01/13/25 17:13		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134256

Review By	Sarabjit Jaswal	Review On	1/13/2025 6:24:27 PM
Supervise By	Mohan Bera	Supervise On	1/13/2025 6:25:37 PM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	MP83506
Chk Standard	MP83513,MP83514

19	PB165922BS	LCS922	LCS	01/13/25 17:17		Kareem	OK
20	CCV148	CCV148	CCV	01/13/25 17:21		Kareem	OK
21	CCB148	CCB148	CCB	01/13/25 17:24		Kareem	OK