

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51879MA No.: _____ SDG No.: MBHMQ0Instrument ID: P5Start Date : 01/08/2025Analytical Method: ICP-AESRun Batch : LB134200

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999545	0.092286	-0.001072	WLR Blank	Inverse Conc
Antimony	0.999572	0.104277	-0.000141	WLR Blank	Inverse Conc
Arsenic	0.999748	169.200700	1.649988	WLR Blank	Inverse Conc
Barium	0.999963	3.475095	0.007272	WLR Blank	Inverse Conc
Beryllium	0.999878	10.980928	-0.000771	WLR Blank	Inverse Conc
Boron	0.999316	1.435209	0.010928	WLR Blank	Inverse Conc
Cadmium	0.999860	3.669604	0.000297	WLR Blank	Inverse Conc
Calcium	0.999804	0.053466	0.009646	WLR Blank	Inverse Conc
Chromium	0.999969	0.223953	-0.000242	WLR Blank	Inverse Conc
Cobalt	0.999992	0.750356	-0.000865	WLR Blank	Inverse Conc
Copper	0.999551	7.079540	0.081833	WLR Blank	Inverse Conc
Iron	0.999979	0.033519	-0.000205	WLR Blank	Inverse Conc
Lead	0.999828	0.299280	-0.001607	WLR Blank	Inverse Conc
Lithium	0.999485	1.283855	0.070345	WLR Blank	Inverse Conc
Magnesium	0.999880	0.009285	-0.001167	WLR Blank	Inverse Conc
Manganese	0.999897	0.347463	-0.001402	WLR Blank	Inverse Conc
Molybdenum	0.999777	0.444697	0.000967	WLR Blank	Inverse Conc
Nickel	0.999910	0.663165	-0.000727	WLR Blank	Inverse Conc
Phosphorus	0.999272	0.077934	-0.001343	WLR Blank	Inverse Conc
Potassium	0.999913	0.020138	0.240910	WLR Blank	Inverse Conc
Selenium	0.999569	0.056761	-0.000524	WLR Blank	Inverse Conc
Silicon	0.953238	1.316477	0.020496	WLR Blank	Inverse Conc
Silver	0.999969	0.258167	0.000739	WLR Blank	Inverse Conc
Sodium	0.999290	0.927156	-0.011061	WLR Blank	Inverse Conc
Strontium	0.999948	9.004102	0.000437	WLR Blank	Inverse Conc
Sulfur	0.999068	0.037338	0.001301	WLR Blank	Inverse Conc
Thallium	0.999938	0.100990	-0.000547	WLR Blank	Inverse Conc
Tin	0.999817	0.189996	-0.000145	WLR Blank	Inverse Conc
Vanadium	0.999784	7.036516	-0.009138	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHMQ0
Instrument ID: P5 Start Date : 01/08/2025
Analytical Method: ICP-AES Run Batch : LB134200

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Zinc	0.999815	1.983433	-0.000192	WLR Blank	Inverse Conc

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHMQ0
Instrument ID: P5 Start Date : 01/08/2025
Analytical Method: ICP-AES Run Batch : LB134200

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------