

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134241

Review By	Mohan Bera	Review On	1/12/2025 8:53:42 PM
Supervise By	Sarabjit Jaswal	Supervise On	1/12/2025 8:54:07 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501		
ICV Standard	MP83508		
CCV Standard	MP83511		
ICSA Standard	MP83509,MP83510		
CRI Standard			
LCS Standard			
Chk Standard	MP83513,MP83514		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	01/10/25 11:09		Kareem	OK
2	S1	S01	CAL2	01/10/25 11:13		Kareem	OK
3	S2	S02	CAL3	01/10/25 11:17		Kareem	OK
4	S3	S03	CAL4	01/10/25 11:21		Kareem	OK
5	S4	S04	CAL5	01/10/25 11:25		Kareem	OK
6	S5	S05	CAL6	01/10/25 11:28		Kareem	OK
7	S6	S06	CAL7	01/10/25 11:32		Kareem	OK
8	ICV057	ICV057	ICV	01/10/25 13:45	Inst. paused for calibration review	Kareem	OK
9	ICB057	ICB057	ICB	01/10/25 13:50		Kareem	OK
10	ICSA057	ICSA057	ICSA	01/10/25 13:54		Kareem	OK
11	ICSAB057	ICSAB057	ICSAB	01/10/25 14:27	Inst. paused for review	Kareem	OK
12	CCV144	CCV144	CCV	01/10/25 14:41	Inst. paused for review	Kareem	OK
13	CCB144	CCB144	CCB	01/10/25 14:45		Kareem	OK
14	P5203-06	MBHN32	SAM	01/10/25 14:49		Kareem	OK
15	P5203-07	MBHN33	SAM	01/10/25 14:53		Kareem	OK
16	P5203-08	MBHN61	SAM	01/10/25 14:56		Kareem	OK
17	P5203-09	MBHN62	SAM	01/10/25 15:00		Kareem	OK
18	P5203-10	MBHN63	SAM	01/10/25 15:04		Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134241

Review By	Mohan Bera	Review On	1/12/2025 8:53:42 PM
Supervise By	Sarabjit Jaswal	Supervise On	1/12/2025 8:54:07 PM

STD. NAME	STD REF.#
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501
ICV Standard	MP83508
CCV Standard	MP83511
ICSA Standard	MP83509,MP83510
CRI Standard	
LCS Standard	
Chk Standard	MP83513,MP83514

19	P5203-11	MBHN95	SAM	01/10/25 15:22		Kareem	OK
20	P5203-12	MBHNB5	SAM	01/10/25 15:26		Kareem	OK
21	P5203-14	MBHNB7	SAM	01/10/25 15:34		Kareem	OK
22	P5203-15	MBHNB8	SAM	01/10/25 15:37		Kareem	OK
23	P5203-16	MBHNB9	SAM	01/10/25 15:41		Kareem	OK
24	P5203-17	MBHNC0	SAM	01/10/25 15:45		Kareem	OK
25	P5203-21	MBHND0	SAM	01/10/25 15:49		Kareem	OK
26	P5202-10DL	MBHNG5	SAM	01/10/25 15:52	2x for Ca	Kareem	Confirms
27	P5202-11DL	MBHNG5D	DUP	01/10/25 15:56	2x for Ca	Kareem	Confirms
28	P5202-10LDL	MBHNG5L	SD	01/10/25 16:00	10x for Ca	Kareem	Confirms
29	P5202-12DL	MBHNG5S	MS	01/10/25 16:04	2x for Ca	Kareem	Confirms
30	P5207-12DL	MBHN38	SAM	01/10/25 16:08	2x for Ca	Kareem	Confirms
31	P5207-22DL	MBHN60	SAM	01/10/25 16:11	2x for Ca	Kareem	Confirms
32	P5198-09DL	MBHMP2	SAM	01/10/25 16:15	2x for Ca	Kareem	Confirms
33	CCV145	CCV145	CCV	01/10/25 16:19		Kareem	OK
34	CCB145	CCB145	CCB	01/10/25 16:22		Kareem	OK
35	P5198-12DL	MBHMP5	SAM	01/10/25 16:26	5x for Mn	Kareem	Confirms
36	P5198-05A	MBHMP0A	PS	01/10/25 16:30	PS for As,Se,Tl	Kareem	OK
37	P5202-10A	MBHNG5A	PS	01/10/25 16:34	PS for Zn, 0.055 - M5801-10 ml sample	Kareem	OK

Instrument ID: P5

Daily Analysis Runlog For Sequence/QC Batch ID # LB134241

Review By	Mohan Bera	Review On	1/12/2025 8:53:42 PM
Supervise By	Sarabjit Jaswal	Supervise On	1/12/2025 8:54:07 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83500,MP83507,MP83505,MP83504,MP83503,MP83502,MP83501		
ICV Standard	MP83508		
CCV Standard	MP83511		
ICSA Standard	MP83509,MP83510		
CRI Standard			
LCS Standard			
Chk Standard	MP83513,MP83514		

38	P5203-18A	MBHNC1A	PS	01/10/25 16:38	NOT USE	Kareem	Not Ok
39	P5207-13A	MBHN39A	PS	01/10/25 16:42	PS for Cu,Tl	Kareem	OK
40	P5198-17DL	MBHMMW4	SAM	01/10/25 16:45	5x for Ca	Kareem	Confirms
41	P5203-13	MBHNB6	SAM	01/10/25 17:13		Kareem	OK
42	CCV146	CCV146	CCV	01/10/25 17:19		Kareem	OK
43	CCB146	CCB146	CCB	01/10/25 17:23		Kareem	OK