

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLCContract: 68HERH20D0011Lab Code: ACE Case No.: 51879MA No.: _____ SDG No.: MBHN53Instrument ID: P4Start Date : 01/12/2025Analytical Method: ICP-AESRun Batch : LB134242

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
Aluminum	0.999630	0.013481	0.000102	Lin. Reg	NONE
Antimony	0.997818	0.175170	0.000425	Lin. Reg	NONE
Arsenic	0.997932	0.092726	-0.000036	Lin. Reg	NONE
Barium	0.999858	0.457506	0.013631	Lin. Reg	NONE
Beryllium	0.999734	5.831456	0.000087	Lin. Reg	NONE
Cadmium	0.999846	3.680984	0.000026	Lin. Reg	NONE
Calcium	0.999773	0.009832	0.000016	Lin. Reg	NONE
Chromium	0.999698	0.200102	0.000041	Lin. Reg	NONE
Cobalt	0.999933	1.052193	-0.000050	Lin. Reg	NONE
Copper	0.999924	0.241369	0.000772	Lin. Reg	NONE
Iron	0.999995	0.006004	0.000001	Lin. Reg	NONE
Lead	0.999954	0.288689	-0.000148	Lin. Reg	NONE
Magnesium	0.999702	0.003437	-0.000067	Lin. Reg	NONE
Manganese	0.999985	0.085459	0.000012	Lin. Reg	NONE
Nickel	0.999875	0.747494	-0.000039	Lin. Reg	NONE
Phosphorus	0.999810	0.066421	-0.001713	Lin. Reg	NONE
Potassium	0.999054	0.004293	0.007528	Lin. Reg	NONE
Selenium	0.998404	0.097877	0.000407	Lin. Reg	NONE
Silver	0.999756	0.328618	-0.000377	Lin. Reg	NONE
Sodium	0.999782	0.006409	-0.002939	Lin. Reg	NONE
Thallium	0.999576	0.120880	0.000180	Lin. Reg	NONE
Tin	0.999924	0.203386	0.000428	Lin. Reg	NONE
Titanium	0.999795	0.542944	0.003021	Lin. Reg	NONE
Vanadium	0.999828	0.189522	0.000567	Lin. Reg	NONE
Zinc	0.996413	2.380133	0.002431	Lin. Reg	NONE

FORM 16-IN

INITIAL CALIBRATION SUMMARY

Lab Name: Alliance Technical Group, LLC Contract: 68HERH20D0011
Lab Code: ACE Case No.: 51879 MA No.: SDG No.: MBHN53
Instrument ID: P4 Start Date : 01/12/2025
Analytical Method: ICP-AES Run Batch : LB134242

Analyte	Corr. Coeff.	Slope	Intercept	Calib. Type	Weighting
---------	--------------	-------	-----------	-------------	-----------