

Order ID:

P5213

284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789

8900, Fax: 908 789 8922

Prep Standard - Chemical Standard Summary

Test :	Percent Solids,pH,TOC
Prepbatch ID :	
Sequence ID/Qc Bate	ch ID: LB133854,LB133969,
Standard ID : WP109217,WP10921	18,WP109225,WP110662,WP110663,WP110664,WP110665,WP110666,WP110667,
Chemical ID : 1111111,111112,W2784	4,W2860,W3005,W3071,W3072,W3093,W3094,W3107,W3111,W3112,



Alliance

Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	<u>NAME</u>	NO.	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Mohan Bera	
2050	CALE_5 (WC IPETTE_3 08/16/2024								
FROM 5.00000ml of W2860 + 8.51200gram of W3111 + 990.00000ml of W3112 = Final Quantity: 1000.000 ml									

Recipe				Expiration	<u>Prepared</u>			Supervised By
<u>ID</u>	<u>NAME</u>	<u>NO.</u>	Prep Date	<u>Date</u>	<u>By</u>	<u>ScaleID</u>	<u>PipetteID</u>	Mohan Bera
2051	TOC STOCK STD-SS, 4000PPM	WP109218	08/07/2024	02/07/2025	Iwona Zarych	WETCHEM_S	WETCHEM_F	
						CALE_5 (WC	IPETTE_3	08/16/2024

FROM 5.00000ml of W2860 + 8.51200gram of W2784 + 990.00000ml of W3112 = Final Quantity: 1000.000 ml



Alliance

Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	<u>NAME</u>	<u>NO.</u>	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Mohan Bera
2435	1:1 PHOSPHORIC ACID FOR TOC SOILS	WP109225	08/07/2024	02/07/2025	lwona Zarych	None	WETCHEM_F IPETTE_3	08/16/2024
	50 00000 L (MO000 - 50 00000 L	5340440	E: 10 ::				(WC)	

FROM 50.00000ml of W2860 + 50.00000ml of W3112 = Final Quantity: 100.000 ml

Recipe ID	NAME	<u>NO.</u>	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Iwona Zarych
712	TOC SOIL cal 250ppm	<u>WP110662</u>	11/12/2024	11/19/2024	Niha Farheen Shaik	None	None	11/14/2024

FROM 15.00000ml of W3112 + 1.00000ml of WP109217 = Final Quantity: 16.000 ml





Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

710 TOC SOIL cal 500ppm WP110663 11/12/2024 11/19/2024 Niha Farheen Shaik None None 11/14/2024	Recipe ID	NAME	<u>NO.</u>	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipetteID</u>	Supervised By Iwona Zarych
	710	TOC SOIL cal 500ppm	<u>WP110663</u>	11/12/2024	11/19/2024		None	None	,

FROM	14.00000ml of W3112 + 2.00000ml of WP109217 = Final Quantity: 16.000 ml
------	---

Recipe				Expiration	Prepared			Supervised By
<u>ID</u>	<u>NAME</u>	<u>NO.</u>	Prep Date	<u>Date</u>	<u>By</u>	<u>ScaleID</u>	<u>PipetteID</u>	Iwona Zarych
3544	TOC SOIL Cal- CCV 1000PPM	WP110664	11/12/2024	11/19/2024	Niha Farheen	None	None	
					Shaik			11/14/2024

FROM 15.00000ml of W3112 + 5.00000ml of WP109217 = Final Quantity: 20.000 ml



Alliance

Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID	NAME	<u>NO.</u>	Prep Date	Expiration Date	Prepared By	<u>ScaleID</u>	<u>PipettelD</u>	Supervised By Iwona Zarych
713	TOC SOIL cal 2000ppm	<u>WP110665</u>	11/12/2024	11/19/2024	Niha Farheen Shaik	None	None	11/14/2024

Recipe				Expiration	<u>Prepared</u>			Supervised By
<u>ID</u>	NAME	NO.	Prep Date	<u>Date</u>	<u>By</u>	<u>ScaleID</u>	<u>PipetteID</u>	Iwona Zarych
2819	TOC ICV-LCSS, 1000PPM	WP110666	11/12/2024	11/19/2024	Niha Farheen	None	None	
					Shaik			11/14/2024

FROM 15.00000ml of W3112 + 5.00000ml of WP109218 = Final Quantity: 20.000 ml





Fax: 908 789 8922

Wet Chemistry STANDARD PREPARATION LOG

Recipe ID 304	NAME TOC CAL 0.00ppm	<u>NO.</u> WP110667	Prep Date 11/12/2024	Expiration Date 11/19/2024	Prepared By Niha Farheen Shaik	ScaleID None	PipetteID None	Supervised By Iwona Zarych 11/14/2024
FROM	100.00000ml of W3112 = Final Quar	ntity: 100.00	0 ml					



CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	P243-500 / Potassium Hydrogen Phthalate, 500 gms	201089	06/30/2025	12/23/2020 / apatel	12/16/2020 / apatel	W2784
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	J0260-3 / Phosphoric Acid, 2.5 L	0000278313	01/31/2026	07/12/2021 / apatel	07/12/2021 / apatel	W2860
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL13850-1 / Buffer Solution, PH2 (500ml)	4212E45	12/31/2024	01/31/2023 / lwona	01/31/2023 / Iwona	W3005
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14455-3 / buffer solution pH 7 yellow	4308H30	07/31/2025	01/02/2024 / JIGNESH	12/06/2023 / Iwona	W3071
Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific	AL14940-1 / Buffer	2310P21	04/30/2025	01/02/2024 / JIGNESH	12/07/2023 /	W3072
Supply, Inc.	Solution, PH12 (500ml)			JIGNESH	lwona	
Supply, Inc. Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date /	Chemtech Lot #



CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	1601-1 / PH 10.01 BUFFER,COLOR CD 475ML	4310g83	03/31/2025	04/03/2024 / jignesh	04/02/2024 / jignesh	W3094

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	AL14055-3 / PH 4 BUFFER SOLUTION	AL14055-3	02/27/2026	09/05/2024 / jignesh	05/13/2024 / jignesh	W3107

mCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
00 / Potassium en Phthalate, 500	24A1956910	01/18/2025	06/26/2024 / Iwona	06/26/2024 / Iwona	W3111
)	0 / Potassium	0 / Potassium 24A1956910	0 / Potassium 24A1956910 01/18/2025	0 / Potassium 24A1956910 01/18/2025 06/26/2024 /	Date Opened By Received By 0 / Potassium 24A1956910 01/18/2025 06/26/2024 / 06/26/2024 /

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / Iwona	07/03/2024 / Iwona	W3112

Phosphoric Acid BAKER ANALYZED® A.C.S. Reagent

(orthophosphoric acid)



Material No.: 0260-03 Batch No.: 0000278313 Manufactured Date: 2021/02/01

Retest Date: 2026/01/31

Revision No: 2

Certificate of Analysis

Meets ACS Reagent Chemical Requirements,

Test	Specification	Result
Assay (H3PO4) (by acidimetry)	85.0 - 87.0 %	85.8
Calcium (Ca)	<= 0.002 %	< 0.001
Color (APHA)	<= 10	5
Insoluble Matter	<= 0.001 %	< 0.001
ACS – Magnesium (Mg)	<= 0.002 %	< 0.002
Sulfate (SO4)	<= 12 ppm	< 4
Volatile Acids (as CH₃COOH)	<= 0.001 %	0.001
Reducing Substances	Passes Test	PT
Chloride (Cl)	<= 3 ppm	< 1
Nitrate (NO3)	<= 5 ppm	< 2
Trace Impurities - Antimony (Sb)	<= 20.000 ppm	0.007
Trace Impurities – Arsenic (As)	<= 0.500 ppm	< 0.001
Trace Impurities – Iron (Fe)	<= 10.000 ppm	< 1.000
Heavy Metals (as Pb)	<= 8 ppm	< 3
Frace Impurities – Manganese (Mn)	<= 0.500 ppm	0.005
Trace Impurities – Potassium (K)	<= 40.000 ppm	< 0.001
Trace Impurities – Sodium (Na)	<= 200.000 ppm	0.082

For Laboratory, Research or Manufacturing Use

Exceeds A.C.S. Specifications

Meets Reagent Specifications for testing USP/NF monographs

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC





RICCA CHEMICAL COMPANY®

O.

1490 Lammers Pike Batesville, IN 47006 http://www.riccachemical.com

1-888-GO-RICCA customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4308H30

Product Number: 1551

Manufacture Date: AUG 09, 2023

Expiration Date: JUL 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

5 10 15 20 25 35 40 45 Hq 7.12 7.09 7.06 7.04 7.027.00 6.99 6.98 6.98 6.97 6.97

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Phosphate Dibasic	7558-79-4	ACS
Potassium Dihydrogen Phosphate	7778-77-0	ACS
Preservative	Proprietary	
Yellow Dye	Proprietary	coccottiti S. Tues and et e e e
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result	
Appearance	Yellow liquid	Passed	*Not a certified value
Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.002	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer A	ASTM (D 5464)
Buffer A	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months
Possesses de J. Character 1500	***************************************	24 months

Youl Drandon

Paul Brandon (08/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 4308H30 Product Number: 1551 Page 2 of 2



RICCA CHEMICAL COMPANY®

W 3072

MC. (2/01/23)

Certificate of Analysis

1841 Broad Street Pocomoke City, MD 21851 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

Buffer, Reference Standard, pH 12.00 ± 0.01 at 25°C

Lot Number: 2310P21

Product Number: 1615

Manufacture Date: OCT 24, 2023

Expiration Date: APR 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

°C 15 35 40 12.35 12.17 11.99 11.78 Hg 11.62

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Sodium Hydroxide	1310-73-2	Reagent

Test	Specification	Result	
Appearance	Colorless liquid	Passed	*Not a certified value.
		· · · · · · · · · · · · · · · · · · ·	

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	12.005	0.02	186-I-g, 186-II-g, 191d

pH measurements were performed in our Pocomoke City, MD laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.01) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1615-1	4 L natural poly	18 months
1615-16	500 mL clear PET-G	18 months
1615-32	1 L natural poly	18 months
1615-5	20 L Cubitainer®	18 months

Storen Travers.

Sharon Travers (10/24/2023)

Operations Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 2310P21 Product Number: 1615 Page 2 of 2



Certificate of Analysis

1 Reagent Lane Fair Lawn, NJ 07410 201.796.7100 tel 201.796.1329 fax

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2015 by SAI Global Certificate Number CERT – 0120632

This is to certify that units of the lot number below were tested and found to comply with the specifications of the grade listed. Certain data have been supplied by third parties. Thermo Fisher Scientific expressly disclaims all warranties, expressed or implied, including the implied warranties of merchantability and fitness for a particular purpose. Products are for research use or further manufacturing. Not for direct administration to humans or animals. It is the responsibility of the final formulator and end user to determine suitability based upon the intended use of the end product. Products are tested to meet the analytical requirements of the noted grade. The following information is the actual analytical results obtained.

Catalog Number	P243	Quality Test / Release Date	06/19/2020
Lot Number	201089	•	
Description	POTASSIUM HYDROGEN PHTHALATE	ACIDIMETRIC STANDARD, A.C.S	S.
Country of Origin	Spain	Suggested Retest Date	Jun/2025
Chemical Origin	Organic - non animal		
BSE/TSE Comment	No animal products are used as starting r processing aids, or any other material that		

N/A			
Result Name	Units	Specifications	Test Value
APPEARANCE		REPORT	WHITE CRYSTALS
ASSAY POTASSIUM HYDROGEN PHTHALATE	%	Inclusive Between 99.95 - 100.05	100.03
CHLORINE COMPOUNDS	%	<= 0.003	<0.003
HEAVY METALS (as Pb)	ppm	<= 5	<5
IDENTIFICATION	PASS/FAIL	= PASS TEST	PASS TEST
INSOLUBLE MATTER	%	<= 0.005	<0.005
IRON (Fe)	ppm	<= 5	<5
PH OF 0.05M SOLUTION		Inclusive Between 4.00 - 4.02	4.00
SODIUM (Na)	%	<= 0.005	<0.005
SULFUR COMPOUNDS	%	<= 0.002	<0.002%
TRACEABLE TO NIST	SOD CARBONATE	= LOT 351a	351a
TRACEABLE TO NIST KHP STD	POT. ACID PHTHALATE	= LOT 84L	84L

Julian Burton

Julian Burton - Quality Control Manager - Fair Lawn

^{*}Based on suggested storage condition.



W 3005 Mec. 1/31/23

1490 Lammers Pike Batesville, IN 47006 http://www.riccachemical.com 1-888-GO-RICCA customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 2.00 ± 0.01 at 25° C

Lot Number: 4212E45

Product Number: 1493

Manufacture Date: DEC 20, 2022

Expiration Date: DEC 2024

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ±0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05.

°C 10 15 20 25 30 35 40 45 50 pН 1.93 1.98 1.98 2.00 2.01 2.03 2.03 2.04 2.04

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Potassium Chloride	7447-40-7	ACS
Hydrochloric Acid	7647-01-0	ACS

Test	Specification	Result	
Appearance	Colorless liquid	Passed	*Not a certified value.
Test	A		

Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	2.000	0.02	185i, 186-I-g, 186-II-g
**	***********************		1001' 100 T.S' 100-II-B

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1493-1	4 L natural poly	24 months
1493-16	500 mL natural poly	24 months
1493-32	1 L natural poly	24 months
1493-5	20 L Cubitainer®	24 months

faul Drandon

Paul Brandon (12/20/2022)

Production Manager

This Certificate of Analysis is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 4212E45 Product Number: 1493 Page 2 of 2



RICCA CHEMICAL COMPANY

1490 Lammers Pike Batesville, IN 47006 http://www.riccachemical.com 1-888-GO-RICCA

customerservice@riccachemical.com

Certificate of Analysis Onlong Concession Co

Buffer, Reference Standard, pH 7.00 ± 0.01 at 25°C (Color Coded Yellow)

Lot Number: 4401F99

Product Number: 1551

Manufacture Date: JAN 08, 2024

Expiration Date: DEC 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

5 10 15 20 25 30 35 40 45 50 pН 7.12 7.09 7.06 7.04 7.02 7.00 6.99 6.98 6.98 6.97 6.97

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/EP	
Sodium Phosphate Dibasic	7558-79-4	ACS	
Potassium Dihydrogen Phosphate	7778-77-0	ACS	
Preservative	Proprietary	11.11.77	
Yellow Dye	Proprietary		
Sodium Hydroxide	1310-73-2		

Test	Specification	Result	
Appearance	Yellow liquid	Passed	*Not a certified value
Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	7.004	0.02	186-I-g, 186-II-g, 191d

Specification	Reference	
Commercial Buffer Solutions	ASTM (D 1293 B)	
Buffer A	ASTM (D 5464)	
Buffer A	ASTM (D 5128)	

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1551-1	4 L natural poly	24 months
1551-1CT	4 L Cubitainer®	24 months
1551-2.5	10 L Cubitainer®	24 months
1551-5	20 L Cubitainer®	24 months
		V V V I I 111 155 5 1

faul Drandon

Paul Brandon (01/08/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 4401F99 Product Number: 1551 Page 2 of 2



RICCA CHEMICAL COMPANY®

1490 Lammers Pike Batesville, IN 47006 http://www.riccachemical.com

1-888-GO-RICCA customerservice@riccachemical.com

Certificate of Analysis

Buffer, Reference Standard, pH 10.00 ± 0.01 at 25°C (Color Coded Blue)

Lot Number: 4310G83

Product Number: 1601

Manufacture Date: OCT 09, 2023

Expiration Date: MAR 2025

The certified value for this product is confirmed in independent testing by a second qualified chemist. The NIST traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

°C 0 5 10 15 20 25 30 35 40 50

pH 10.31 10.23 10.17 10.11 10.05 10.00 9.95 9.91 9.87 9.81

Name	CAS#	Grade
Water	7732-18-5	ACS/ASTM/USP/EP
Sodium Carbonate	497-19-8	ACS
Sodium Bicarbonate	144-55-8	ACS
Sodium Hydroxide	1310-73-2	Reagent
Preservative	Proprietary	2008000
Blue Dye	Proprietary	
		De la companya del companya de la companya del companya de la comp

Test	Specification	Result	
Appearance	Blue liquid	Passed	*Not a certified value.
Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	10.003	0.02	186-I-g, 186-II-g, 191d

Specification	Reference
Commercial Buffer Solutions	ASTM (D 1293 B)
Buffer C	ASTM (D 5464)
Buffer C	ASTM (D 5128)

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)
1601-16	500 mL natural poly	18 months
1601-5	20 L Cubitainer®	18 months

Hand Brandon

Paul Brandon (10/09/2023)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 4310G83 Product Number: 1601 Page 2 of 2



RICCA CHEMICAL COMPANY

customerservice@riccachemical.com

1490 Lammers Pike Batesville, IN 47006 http://www.riccachemical.com

1-888-GO-RICCA

Certificate of Analysis

Manufacture Date: MAR 09, 2024

Expiration Date: FEB 2026

Buffer, Reference Standard, pH 4.00 ± 0.01 at 25°C (Color Coded Red)

Lot Number: 4403F90

Product Number: 1501

The certified value for this product is confirmed in independent testing by a second qualified chemist.

The NIST Traceable pH value is certified to ± 0.01 at 25 °C only. All other pH values at their corresponding temperatures are accurate to ± 0.05 .

10 15 20 25 30 35 45 50 4.00 4.00 pН 4.00 4.00 4.00 4.00 4.01 4.02 4.03 4.04 4.06

Name	CAS#	Grade	
Water	7732-18-5	ACS/ASTM/USP/I	EP
Potassium Acid Phthalate	877-24-7	Buffer	
Preservative	Proprietary	Commercial	• •
Red Dye	Proprietary	Purified	
Test	Specification	Result	STATE OF STATE OF
Appearance	Red liquid	Passed	*Not a certified value
Test	Certified Value	Uncertainty	NIST SRM#
pH at 25°C (Method: SQCP027, SQCP033)	4.000	0.02	185i, 186-I-g, 186-II-g

Specification	Reference	
Commercial Buffer Solutions	ASTM (D 1293 B)	
Buffer B	ASTM (D 5464)	
Buffer B	ASTM (D 5128)	

pH measurements were performed in our Batesville, IN laboratory under ISO/IEC 17025 accreditation (ANAB Certificate L2387.02) and are certified traceable to National Institute of Standards and Technology (NIST) Standard Reference Material as indicated above via an unbroken chain of comparisons. The uncertainty is calculated from the uncertainty of the measurement variation from sample to sample, the uncertainty in the NIST Standard Reference Material, and the uncertainty of the measurement process. The uncertainty is multiplied by k=2, corresponding to 95% coverage in a normal distribution. Volumetric glassware complies with Class A tolerance requirements of ASTM E 288 and NIST Circular 434; it is calibrated before first use and recalibrated regularly in accordance with ASTM E 542 and NIST Procedure NBSIR 74-461. Balances are calibrated regularly with weights certified traceable to the NIST national mass standard. Thermometers and temperature probes are calibrated before first use and recalibrated regularly with a thermometer traceable to NIST standards. All products are prepared according to master documents that assure manufacture according to validated methods. Batch records document raw material traceability and production and testing history for each lot manufactured.

Part Number	Size / Package Type	Shelf Life (Unopened Container)	
1501-2.5	10 L Cubitainer®	24 months	
1501-32	1 L natural poly	24 months	
1501-5	20 L Cubitainer®	24 months	

Hand Brandon

Paul Brandon (03/09/2024)

Production Manager

This document is designed to comply with ISO Guide 31 "Reference Materials -- Contents of Certificates and Labels."

This product was tested in an ISO 17025 Accredited Laboratory

This test report shall not be reproduced, except in full, without the written approval of Ricca Chemical Company.

Version: 1.3 Lot Number: 4403F90 Product Number: 1501 Page 2 of 2



Certificate of Analysis

01/19/2022

01/18/2025

POTASSIUM HYDROGEN PHTHALATE

Material: N983

Grade: ACS GRADE Batch Number: 24A1956910

Chemical Formula: HOOCC6H4COOK

Molecular Weight: 204.22

CAS #: 877-24-7

Appearance: Storage: Room Temperature

White crystals.

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Assay (dried basis)	99.95 - 100.05 %	99.97 %	PASS
Chlorine Compounds	<= 0.003 %	<0.003 %	PASS
Heavy Metals (as Pb)	<= 5 ppm	<5 ppm	PASS
Insoluble Matter	<= 0.005 %	0.003 %	PASS
Iron	<= 5 ppm	<5 ppm	PASS
pH (0.05M, Water) @25C	4.00 - 4.02	4.00	PASS
Sodium	<= 0.005 %	<0.005 %	PASS
Sulfur Compounds	<= 0.002 %	<0.002 %	PASS

Manufacture Date:

Reassay Date:

Spec Set: N983ACS

Internal ID #: 710

Signature Additional Information

We certify that this batch conforms to the specifications listed.

This document has been electronically produced and is valid

without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon

VWR Chemicals, LLC.

28600 Fountain Parkway, Solon OH 44139 USA

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.