

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB133854

Slope : 98.6

BalanceID: WC SC-7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER, COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	12/10/2024	08:00
2	CAL2	1	Water	NA	NA	20.2	7.00	12/10/2024	08:02
3	CAL3	1	Water	NA	NA	20.2	10.02	12/10/2024	08:03
4	ICV	1	Water	NA	NA	20.3	7.00	12/10/2024	08:05
5	CCV1	1	Water	NA	NA	20.3	2.01	12/10/2024	08:10
6	P5213-01	1	Solid	20.02	20	23.1	5.00	12/10/2024	08:15
7	P5220-01	1	Solid	20.03	20	23.6	5.72	12/10/2024	08:22
8	P5220-01DUP	1	Solid	20.04	20	23.7	5.73	12/10/2024	08:23
9	CCV2	1	Water	NA	NA	20.3	12.02	12/10/2024	08:25

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s p5213

WorkList ID : 186158

Department : Wet-Chemistry

Date : 12-10-2024 07:46:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P5213-01	TAPIAL3-SB031-15-120624-00-	Solid	pH	Cool 4 deg C	WEST04	L51	12/06/2024	9045D
P5220-01	GAS-AUD-1562	Solid	pH	Cool 4 deg C	PSEG03	L41	12/09/2024	9045D
P5222-01	TRE-1113	Solid	pH	Cool 4 deg C	PSEG03	L51	12/09/2024	9045D

12

12110/2024

78

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12-10-24 07:50

12-10-24

12-10-24

Date/Time

Raw Sample Received by:

Raw Sample Relinquished by:

12-10-24

12-10-24

12-10-24