



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5213

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133854

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV						
pH		pH	7.00	7	100	90-110	12/10/2024
Sample ID:	CCV1						
pH		pH	2.01	2.00	101	90-110	12/10/2024
Sample ID:	CCV2						
pH		pH	12.02	12.00	100	90-110	12/10/2024

Initial and Continuing Calibration Verification

Client: Weston Solutions

SDG No.: P5213

Project: Ft Meade Tipton Airfield Parcel RI - PO 0111169

RunNo.: LB133969

Analyte		Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID:	ICV1						
TOC		mg/L	934	1000	93	90-110	11/12/2024
Sample ID:	CCV1						
TOC		mg/L	986	1000	99	90-110	12/17/2024
Sample ID:	CCV2						
TOC		mg/L	1060	1000	106	90-110	12/17/2024