


LB133887

7471B

INSTRUMENT ID: CV1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.0	0.000	0.063	0.063	511	0.000	511					1
0.2	0.200	0.219	0.019	1290	0.0 %	1290					9
2.5	2.500	2.463	-0.037	12479	0.0 %	12479					1
5.0	5.000	4.860	-0.140	24437	0.0 %	24437					3
7.5	7.500	7.575	0.075	37976	0.0 %	37976					1
10.0	10.000	10.019	0.019	50168	0.0 %	50168					0

LB133887 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 11 Dec 2024 10:53:39

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	511	-	0.0000	7471B	PPB	11 Dec 2024 10:56:02	S	Std
0.2 - 1	50.2	1290	-	0.2000	7471B	PPB	11 Dec 2024 10:58:19	S	Std
2.5 - 1	52.5	12479	-	2.5000	7471B	PPB	11 Dec 2024 11:00:36	S	Std
5.0 - 1	55	24437	-	5.0000	7471B	PPB	11 Dec 2024 11:02:53	S	Std
7.5 - 1	57.5	37976	-	7.5000	7471B	PPB	11 Dec 2024 11:05:11	S	Std
10.0 - 1	510	50168	-	10.0000	7471B	PPB	11 Dec 2024 11:10:31	S	Std
ICV16 - 1	ICV16	19571	3.8847	-	7471B	PPB	11 Dec 2024 11:13:27	U	SMPL
ICB16 - 1	ICB16	367	0.0343	-	7471B	PPB	11 Dec 2024 11:18:17	U	SMPL
CCV16 - 1	CCV16	24378	4.8485	-	7471B	PPB	11 Dec 2024 11:20:31	U	SMPL
CCB16 - 1	CCB16	8	-0.0377	-	7471B	PPB	11 Dec 2024 11:22:47	U	SMPL
CRA - 1	CRA	1227	0.2067	-	7471B	PPB	11 Dec 2024 11:25:07	U	SMPL
HighStd - 1	HighStd	49178	9.8209	-	7471B	PPB	11 Dec 2024 11:27:22	U	SMPL
ChkStd - 1	ChkStd	34166	6.8110	-	7471B	PPB	11 Dec 2024 11:29:38	U	SMPL
PB165560BL - 1	PB165560BL	273	0.0155	-	7471B	PPB	11 Dec 2024 11:32:00	U	SMPL
PB165560BS - 1	PB165560BS	17973	3.5643	-	7471B	PPB	11 Dec 2024 11:34:19	U	SMPL
P5174-01 - 1	ROLL-OFF-COMP	184418	36.9366	-	7471B	PPB	11 Dec 2024 11:36:35	U	SMPL
P5175-01 - 1	OK-01-12062024	-521	-0.1437	-	7471B	PPB	11 Dec 2024 11:38:54	U	SMPL
P5176-01 - 1	EO-02-12062024	2141	0.3900	-	7471B	PPB	11 Dec 2024 11:41:25	U	SMPL
P5176-03 - 1	EO-03-12062024	3467	0.6559	-	7471B	PPB	11 Dec 2024 11:43:43	U	SMPL
P5196-01 - 1	MH-761	1470	0.2556	-	7471B	PPB	11 Dec 2024 11:46:01	U	SMPL
CCV17 - 1	CCV17	24001	4.7729	-	7471B	PPB	11 Dec 2024 11:48:19	U	SMPL
CCB17 - 1	CCB17	260	0.0128	-	7471B	PPB	11 Dec 2024 11:53:15	U	SMPL
P5213-01 - 1	TAPIAL3-SB03I-15-120624-00-T1	769	0.1149	-	7471B	PPB	11 Dec 2024 11:55:30	U	SMPL
P5216-01 - 1	BP-G-6	5546	1.0727	-	7471B	PPB	11 Dec 2024 11:57:46	U	SMPL
P5216-05 - 1	MH-743	1078	0.1769	-	7471B	PPB	11 Dec 2024 12:00:02	U	SMPL
P5216-09 - 1	MH-744	1172	0.1957	-	7471B	PPB	11 Dec 2024 12:02:20	U	SMPL
P5216-13 - 1	TP-12	4937	0.9506	-	7471B	PPB	11 Dec 2024 12:04:38	U	SMPL
P5219-01 - 1	MR-CAM-09-MR-CAM-010	1645	0.2905	-	7471B	PPB	11 Dec 2024 12:06:54	U	SMPL
P5219-02 - 1	MR-CAM-011	1532	0.2679	-	7471B	PPB	11 Dec 2024 12:09:13	U	SMPL
P5219-03 - 1	MR-CAM-012	2181	0.3980	-	7471B	PPB	11 Dec 2024 12:11:32	U	SMPL
P5219-04 - 1	MR-CAM-013	240	0.0088	-	7471B	PPB	11 Dec 2024 12:13:51	U	SMPL
P5223-01 - 1	TR-05-120924	1760	0.3136	-	7471B	PPB	11 Dec 2024 12:16:08	U	SMPL
CCV18 - 1	CCV18	24220	4.8168	-	7471B	PPB	11 Dec 2024 12:18:24	U	SMPL
CCB18 - 1	CCB18	190	-0.0012	-	7471B	PPB	11 Dec 2024 12:20:40	U	SMPL
P5223-01DUP - 1	TR-05-120924DUP	1666	0.2948	-	7471B	PPB	11 Dec 2024 12:25:47	U	SMPL
P5223-01MS - 1	TR-05-120924MS	19342	3.8388	-	7471B	PPB	11 Dec 2024 12:28:02	U	SMPL
P5223-01MSD - 1	TR-05-120924MSD	18104	3.5906	-	7471B	PPB	11 Dec 2024 12:30:18	U	SMPL
P5236-01 - 1	TAPIAL3-SB03D-3-120724-00-T01	1354	0.2322	-	7471B	PPB	11 Dec 2024 12:32:37	U	SMPL
P5239-01 - 1	WC-1	1358	0.2330	-	7471B	PPB	11 Dec 2024 12:34:57	U	SMPL
P5244-01 - 1	NB-08-121024	2673	0.4977	-	7471B	PPB	11 Dec 2024 12:37:13	U	SMPL
P5245-01 - 1	72-12016	4186	0.8000	-	7471B	PPB	11 Dec 2024 12:39:31	U	SMPL
P5223-01LX5 - 1		244	0.0096	-	7471B	PPB	11 Dec 2024 12:41:47	U	SMPL
P5223-01A - 1		19255	3.8214	-	7471B	PPB	11 Dec 2024 12:44:05	U	SMPL
P5174-01DLX10 - 1		16141	3.1970	-	7471B	PPB	11 Dec 2024 12:46:21	U	SMPL
CCV19 - 1	CCV19	23901	4.7529	-	7471B	PPB	11 Dec 2024 12:51:03	U	SMPL
CCB19 - 1	CCB19	168	-0.0056	-	7471B	PPB	11 Dec 2024 12:53:18	U	SMPL