

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: P5213
Instrument ID: P7
Run Number: LB134024

Contract: WEST04
Sas No.: P5213 SDG No.: P5213
Start Date : 12/19/2024
End Date : 12/19/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Non-Collision Cell									
			Element 165Ho	Q	Element 209Bi	Q	Element	Q	Element	Q	Element	Q
S0	S0	1504	100		100							
S2	S2	1510	102		105							
S3	S3	1514	103		106							
S4	S4	1517	101		104							
S5	S5	1519	97		102							
S6	S6	1522	97		101							
S7	S7	1525	98		100							
S8	S8	1528	97		91							
ICV01	ICV01	1549	101		106							
LLICV	LLICV	1604	99		104							
ICB01	ICB01	1607	102		106							
ICSA01	ICSA01	1610	103		102							
ICSAB01	ICSAB01	1613	104		103							
CCV01	CCV01	1616	102		98							
CCB01	CCB01	1619	101		107							
CRI	CRI	1623	102		107							
PB165717BL	PB165717BL	1626	103		108							
PB165717BS	PB165717BS	1635	102		104							
ZZZZZZ	ZZZZZZ	1639										
P5117-01DUP	TAPIAL3-SB04	1643	99		105							
P5117-01L	TAPIAL3-SB04	1646	101		107							
P5117-01MS	TAPIAL3-SB04	1649	102		107							
P5117-01MSD	TAPIAL3-SB04	1652	103		107							
P5117-01A	TAPIAL3-SB04	1655	102		106							
CCV02	CCV02	1701	99		98							
CCB02	CCB02	1703	98		105							
P5213-01	TAPIAL3-SB03	1707	101		107							
ZZZZZZ	ZZZZZZ	1711										
ZZZZZZ	ZZZZZZ	1714										
CCV03	CCV03	1720	99		97							
CCB03	CCB03	1723	99		107							

FORM 8B

ICP-MS INTERNAL STANDARD RELATIVE INTENSITY SUMMARY

Lab Name: Weston Solutions
Lab Code: CHEM Case no.: P5213
Instrument ID: P7
Run Number: LB134024

Contract: WEST04
Sas No.: P5213 SDG No.: P5213
Start Date : 12/19/2024
End Date : 12/19/2024

Lab SampleID	Client SampleID	Time	Internal Standard %RI For: Collision Cell									
			Element 209Bi	Q	Element	Q	Element	Q	Element	Q	Element	Q
S0	S0	1504	100									
S2	S2	1510	104									
S3	S3	1514	104									
S4	S4	1517	101									
S5	S5	1519	100									
S6	S6	1522	99									
S7	S7	1525	98									
S8	S8	1528	89									
ICV01	ICV01	1549	106									
LLICV	LLICV	1604	105									
ICB01	ICB01	1607	106									
ICSA01	ICSA01	1610	101									
ICSAB01	ICSAB01	1613	100									
CCV01	CCV01	1616	96									
CCB01	CCB01	1619	106									
CRI	CRI	1623	107									
PB165717BL	PB165717BL	1626	107									
PB165717BS	PB165717BS	1635	104									
ZZZZZZ	ZZZZZZ	1639										
P5117-01DUP	TAPIAL3-SB04	1643	106									
P5117-01L	TAPIAL3-SB04	1646	108									
P5117-01MS	TAPIAL3-SB04	1649	107									
P5117-01MSD	TAPIAL3-SB04	1652	106									
P5117-01A	TAPIAL3-SB04	1655	105									
CCV02	CCV02	1701	95									
CCB02	CCB02	1703	106									
P5213-01	TAPIAL3-SB03	1707	110									
ZZZZZZ	ZZZZZZ	1711										
ZZZZZZ	ZZZZZZ	1714										
CCV03	CCV03	1720	94									
CCB03	CCB03	1723	107									