

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133897

Review By	kareem	Review On	12/13/2024 7:07:43 PM
Supervise By	mohan	Supervise On	12/16/2024 11:22:00 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560,MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565,MP83566		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	12/11/24 11:27		Kareem	OK
2	S1	S1	CAL2	12/11/24 11:31		Kareem	OK
3	S2	S2	CAL3	12/11/24 11:35		Kareem	OK
4	S3	S3	CAL4	12/11/24 11:39		Kareem	OK
5	S4	S4	CAL5	12/11/24 11:44		Kareem	OK
6	S5	S5	CAL6	12/11/24 11:48		Kareem	OK
7	ICV01	ICV01	ICV	12/11/24 11:52		Kareem	OK
8	LLICV01	LLICV01	LLICV	12/11/24 12:00		Kareem	OK
9	ICB01	ICB01	ICB	12/11/24 12:39		Kareem	OK
10	CRI01	CRI01	CRDL	12/11/24 12:44		Kareem	OK
11	ICSA01	ICSA01	ICSA	12/11/24 12:48		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	12/11/24 12:52		Kareem	OK
13	CCV01	CCV01	CCV	12/11/24 12:57		Kareem	OK
14	CCB01	CCB01	CCB	12/11/24 13:06		Kareem	OK
15	PB165470TB	PB165470TB	MB	12/11/24 13:10		Kareem	OK
16	PB165527BL	PB165527BL	MB	12/11/24 13:15		Kareem	OK
17	PB165527BS	PB165527BS	LCS	12/11/24 13:19	0.1 ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
18	P5174-02	ROLL-OFF-COMP	SAM	12/11/24 13:23		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133897

Review By	kareem	Review On	12/13/2024 7:07:43 PM
Supervise By	mohan	Supervise On	12/16/2024 11:22:00 PM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

19	P5196-04	MH-761	SAM	12/11/24 13:28		Kareem	OK
20	P5216-04	BP-G-6	SAM	12/11/24 13:32		Kareem	OK
21	P5216-08	MH-743	SAM	12/11/24 13:37		Kareem	OK
22	P5216-12	MH-744	SAM	12/11/24 13:41		Kareem	OK
23	P5216-16	TP-12	SAM	12/11/24 13:46	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
24	P5216-16DUP	TP-12DUP	DUP	12/11/24 13:50	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
25	CCV02	CCV02	CCV	12/11/24 13:54		Kareem	OK
26	CCB02	CCB02	CCB	12/11/24 13:59		Kareem	OK
27	P5245-01	72-12016	SAM	12/11/24 14:08		Kareem	OK
28	CCV03	CCV03	CCV	12/11/24 14:12		Kareem	OK
29	CCB03	CCB03	CCB	12/11/24 14:17		Kareem	OK
30	P5216-16L	TP-12L	SD	12/11/24 14:25	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
31	P5216-16MS	TP-12MS	MS	12/11/24 14:30	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
32	P5216-16MSD	TP-12MSD	MSD	12/11/24 14:34	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
33	P5216-16A	TP-12A	PS	12/11/24 14:38	MS/MSD fail for more than 50% parameters	Kareem	Not Ok
34	PB165528BL	PB165528BL	MB	12/11/24 14:43		Kareem	OK
35	PB165528BS	PB165528BS	LCS	12/11/24 14:47	0.1 ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133897

Review By	kareem	Review On	12/13/2024 7:07:43 PM
Supervise By	mohan	Supervise On	12/16/2024 11:22:00 PM

STD. NAME	STD REF.#
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558
ICV Standard	MP83559
CCV Standard	MP83562
ICSA Standard	MP83560,MP83561
CRI Standard	MP83558
LCS Standard	
Chk Standard	MP83565,MP83566

36	P5216-01	BP-G-6	SAM	12/11/24 14:51		Kareem	OK
37	P5216-05	MH-743	SAM	12/11/24 14:55		Kareem	OK
38	P5216-09	MH-744	SAM	12/11/24 14:59		Kareem	OK
39	P5216-13	TP-12	SAM	12/11/24 15:03		Kareem	OK
40	CCV04	CCV04	CCV	12/11/24 15:07		Kareem	OK
41	CCB04	CCB04	CCB	12/11/24 15:12		Kareem	OK
42	P5217-01	MIXED-DEMO	SAM	12/11/24 15:16		Kareem	OK
43	P5217-01DUP	MIXED-DEMODUP	DUP	12/11/24 15:20		Kareem	OK
44	P5217-01L	MIXED-DEMOL	SD	12/11/24 15:25		Kareem	OK
45	P5217-01MS	MIXED-DEMOMS	MS	12/11/24 15:29	0.1 ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
46	P5217-01MSD	MIXED-DEMOMSD	MSD	12/11/24 15:34	0.1 ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
47	P5217-01A	MIXED-DEMOA	PS	12/11/24 15:38	0.1 ml of m6010 and m6001 were added to 10ml of sample	Kareem	OK
48	P5219-01	MR-CAM-09-MR-CAM	SAM	12/11/24 15:42		Kareem	OK
49	P5219-02	MR-CAM-011	SAM	12/11/24 15:47		Kareem	OK
50	P5219-03DL	MR-CAM-012DL	SAM	12/11/24 15:51	Straight 5x Dilution for all elements	Kareem	OK
51	P5219-04	MR-CAM-013	SAM	12/11/24 15:55	AI high	Kareem	Dilution
52	CCV05	CCV05	CCV	12/11/24 15:59		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB133897

Review By	kareem	Review On	12/13/2024 7:07:43 PM
Supervise By	mohan	Supervise On	12/16/2024 11:22:00 PM
STD. NAME	STD REF.#		
ICAL Standard	MP83552,MP83553,MP83554,MP83555,MP83556,MP83558		
ICV Standard	MP83559		
CCV Standard	MP83562		
ICSA Standard	MP83560,MP83561		
CRI Standard	MP83558		
LCS Standard			
Chk Standard	MP83565,MP83566		

53	CCB05	CCB05	CCB	12/11/24 16:05		Kareem	OK
54	P5219-04DL	MR-CAM-013DL	SAM	12/11/24 16:09	5x for AI	Kareem	Confirms
55	P5223-01	TR-05-120924	SAM	12/11/24 16:14		Kareem	OK
56	CCV06	CCV06	CCV	12/11/24 16:22		Kareem	OK
57	CCB06	CCB06	CCB	12/11/24 16:26		Kareem	OK
58	LR1	LR1	HIGH STD	12/11/24 16:30		Kareem	OK
59	LR2	LR2	HIGH STD	12/11/24 16:35		Kareem	OK
60	CCV07	CCV07	CCV	12/11/24 16:39		Kareem	OK
61	CCB07	CCB07	CCB	12/11/24 16:43		Kareem	OK