

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: EarthEfficient LLC **SDG No.:** P5217
Contract: EART12 **Lab Code:** CHEM **Case No.:** P5217 **SAS No.:** P5217
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Lead	1000	1000	100	90 - 110	P	12/11/2024	11:52	LB133897

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: EarthEfficient LLC **SDG No.:** P5217
Contract: EART12 **Lab Code:** CHEM **Case No.:** P5217 **SAS No.:** P5217
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Lead	10.3	12.0	86	80 - 120	P	12/11/2024	12:00	LB133897

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: EarthEfficient LLC **SDG No.:** P5217
Contract: EART12 **Lab Code:** CHEM **Case No.:** P5217 **SAS No.:** P5217
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Lead	4980	5000	100	90 - 110	P	12/11/2024	12:57	LB133897
CCV02	Lead	5220	5000	104	90 - 110	P	12/11/2024	13:54	LB133897
CCV03	Lead	5070	5000	101	90 - 110	P	12/11/2024	14:12	LB133897
CCV04	Lead	4930	5000	98	90 - 110	P	12/11/2024	15:07	LB133897
CCV05	Lead	4960	5000	99	90 - 110	P	12/11/2024	15:59	LB133897
CCV06	Lead	5130	5000	103	90 - 110	P	12/11/2024	16:22	LB133897
CCV07	Lead	4940	5000	99	90 - 110	P	12/11/2024	16:39	LB133897
