

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: EarthEfficient LLC **SDG No.:** P5217
Contract: EART12 **Lab Code:** CHEM **Case No.:** P5217 **SAS No.:** P5217
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Lead	6.35			-12	12	12/11/2024	12:48	LB133897
ICSAB01	Lead	56.2	49.0	115	37	61	12/11/2024	12:52	LB133897